

SOP ID : M1311-TCLP-15
SDG No : N/A
Weigh By : JP
Balance ID : WC SC-7
pH Meter ID : WC PH METER-1
Extraction By : JP
Filter By : JP
Pippete ID : WC
Tumbler ID : T-1
TCLP Filter ID : 115525

Start Prep Date : 04/14/2025 Time : 16:00
End Prep Date : 04/15/2025 Time : 10:25
Combination Ratio : 20
ZHE Cleaning Batch : N/A
Initial Room Temperature : 23 °C
Final Room Temperature : 22 °C
TCLP Technician Signature : *[Signature]*
Supervisor By : *[Signature]*

Standard Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP110802
HCL-TCLP,1N	N/A	WP110803
HNO3-TCLP,1N	N/A	WP110804
pH Strips	N/A	W1931,W1934,W3171,W3172
pH Strips	W1941,W1942	W3166,W1938,W1939,W1940,
1 Liter Amber	N/A	90424-08
120ml Plastic bottle	N/A	405130101
1:1 HNO3	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

Matrix spikes are added after filtration and before preservation. TUMBLER T-1 checked, 30 rpm. Particle size reduction is not required. Q1808-04 is used for MS-MSD.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/15/25 11:50	<i>[Signature]</i> <i>[Signature]</i> <i>[Signature]</i>	<i>[Signature]</i> <i>[Signature]</i> <i>[Signature]</i>
	Preparation Group	Analysis Group