

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135455

Review By	mohan	Review On	4/17/2025 4:45:45 PM
Supervise By	jaswal	Supervise On	4/17/2025 11:13:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP85245,MP85246,MP85247,MP85248,MP85249,MP85250
ICV Standard	MP85251
CCV Standard	MP85253
ICSA Standard	
CRI Standard	MP85255
LCS Standard	
Chk Standard	MP85252,MP85254,MP85256,MP85258

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	04/16/25 12:56		mohan	OK
2	S0.2	S0.2	CAL2	04/16/25 12:58		mohan	OK
3	S2.5	S2.5	CAL3	04/16/25 13:00		mohan	OK
4	S5	S5	CAL4	04/16/25 13:03		mohan	OK
5	S7.5	S7.5	CAL5	04/16/25 13:05		mohan	OK
6	S10	S10	CAL6	04/16/25 13:10		mohan	OK
7	ICV90	ICV90	ICV	04/16/25 13:13		mohan	OK
8	ICB90	ICB90	ICB	04/16/25 13:15		mohan	OK
9	CCV79	CCV79	CCV	04/16/25 13:17		mohan	OK
10	CCB79	CCB79	CCB	04/16/25 13:20		mohan	OK
11	CRA	CRA	CRDL	04/16/25 13:22		mohan	OK
12	HighStd	HighStd	HIGH STD	04/16/25 13:24		mohan	OK
13	ChkStd	ChkStd	SAM	04/16/25 13:26		mohan	OK
14	PB167607BL	PB167607BL	MB	04/16/25 13:29		mohan	OK
15	PB167607BS	PB167607BS	LCS	04/16/25 13:36		mohan	OK
16	Q1787-12	TP-1	SAM	04/16/25 13:39		mohan	OK
17	Q1788-04	WC-1	SAM	04/16/25 13:41		mohan	OK
18	Q1800-03	WC-A4-01-C	SAM	04/16/25 13:43		mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135455

Review By	mohan	Review On	4/17/2025 4:45:45 PM
Supervise By	jaswal	Supervise On	4/17/2025 11:13:14 PM

STD. NAME	STD REF.#
ICAL Standard	MP85245,MP85246,MP85247,MP85248,MP85249,MP85250
ICV Standard	MP85251
CCV Standard	MP85253
ICSA Standard	
CRI Standard	MP85255
LCS Standard	
Chk Standard	MP85252,MP85254,MP85256,MP85258

19	Q1803-01	WEST-BAY	SAM	04/16/25 13:46		mohan	OK
20	Q1803-02	FUEL-MONITORING	SAM	04/16/25 13:48		mohan	OK
21	CCV80	CCV80	CCV	04/16/25 13:50		mohan	OK
22	CCB80	CCB80	CCB	04/16/25 13:52		mohan	OK
23	Q1808-02	OILY-SOIL-PILE	SAM	04/16/25 13:55		mohan	OK
24	Q1808-04	LAW-25-0060	SAM	04/16/25 13:57		mohan	OK
25	Q1808-04DUP	LAW-25-0060DUP	DUP	04/16/25 13:59		mohan	OK
26	Q1808-04MS	LAW-25-0060MS	MS	04/16/25 14:06		mohan	OK
27	Q1808-04MSD	LAW-25-0060MSD	MSD	04/16/25 14:09		mohan	OK
28	PB167587TB	PB167587TB	MB	04/16/25 14:11		mohan	OK
29	Q1808-04L	LAW-25-0060L	SD	04/16/25 14:13		mohan	OK
30	Q1808-04A	LAW-25-0060A	PS	04/16/25 14:16		mohan	OK
31	CCV81	CCV81	CCV	04/16/25 14:18		mohan	OK
32	CCB81	CCB81	CCB	04/16/25 14:20		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135463

Review By	jaswal	Review On	4/18/2025 11:44:26 AM
Supervise By	Janvi	Supervise On	4/18/2025 12:06:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85022,MP85020,MP85019,MP85018,MP85017		
ICV Standard	MP85023,MP85022		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	04/16/25 12:39		Kareem	OK
2	S1	S1	CAL2	04/16/25 12:44		Kareem	OK
3	S2	S2	CAL3	04/16/25 12:48		Kareem	OK
4	S3	S3	CAL4	04/16/25 12:52		Kareem	OK
5	S4	S4	CAL5	04/16/25 12:56		Kareem	OK
6	S5	S5	CAL6	04/16/25 13:01		Kareem	OK
7	ICV01	ICV01	ICV	04/16/25 13:09		Kareem	OK
8	LLICV1	LLICV1	LLICV	04/16/25 13:37		Kareem	OK
9	ICB01	ICB01	ICB	04/16/25 13:42		Kareem	OK
10	CRI	CRI	CRDL	04/16/25 13:46		Kareem	OK
11	ICSA01	ICSA01	ICSA	04/16/25 13:51		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	04/16/25 13:55		Kareem	OK
13	ICSADL	ICSADL	ICSA	04/16/25 13:59		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	04/16/25 14:03		Kareem	OK
15	CCV01	CCV01	CCV	04/16/25 14:08		Kareem	OK
16	CCB01	CCB01	CCB	04/16/25 14:12		Kareem	OK
17	PB167600BL	PB167600BL	MB	04/16/25 14:38		Kareem	OK
18	PB167600BS	PB167600BS	LCS	04/16/25 14:42		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135463

Review By	jaswal	Review On	4/18/2025 11:44:26 AM
Supervise By	Janvi	Supervise On	4/18/2025 12:06:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85016,MP85022,MP85020,MP85019,MP85018,MP85017
ICV Standard	MP85023,MP85022
CCV Standard	MP85026
ICSA Standard	MP85024,MP85025
CRI Standard	MP85022
LCS Standard	
Chk Standard	MP85030,MP85031

19	Q1806-01	OR-03-041425	SAM	04/16/25 14:46		Kareem	OK
20	Q1807-01	OK-02-041425	SAM	04/16/25 14:50		Kareem	OK
21	Q1807-01L	OK-02-041425L	SD	04/16/25 14:58		Kareem	OK
22	Q1807-01ARE	OK-02-041425A	PS	04/16/25 15:10	Not Required	Kareem	Not Ok
23	Q1808-01	OILY-SOIL-PILE	SAM	04/16/25 15:14		Kareem	OK
24	CCV02	CCV02	CCV	04/16/25 15:18		Kareem	OK
25	CCB02	CCB02	CCB	04/16/25 15:23		Kareem	OK
26	Q1808-03	LAW-25-0060	SAM	04/16/25 15:27		Kareem	OK
27	Q1807-01MS	OK-02-041425MS	MS	04/16/25 15:31		Kareem	OK
28	Q1807-01MSD	OK-02-041425MSD	MSD	04/16/25 15:35		Kareem	OK
29	Q1807-01A	OK-02-041425A	PS	04/16/25 15:41	0.1ML EACH M6004,M6013-10ML SAMPLE	Kareem	OK
30	Q1807-01DUP	OK-02-041425DUP	DUP	04/16/25 15:45		Kareem	OK
31	PB167585BL	PB167585BL	MB	04/16/25 15:49		Kareem	OK
32	PB167585BS	PB167585BS	LCS	04/16/25 15:53		Kareem	OK
33	Q1787-12	TP-1	SAM	04/16/25 15:57		Kareem	OK
34	Q1788-04	WC-1	SAM	04/16/25 16:01		Kareem	OK
35	Q1800-03	WC-A4-01-C	SAM	04/16/25 16:06		Kareem	OK
36	CCV03	CCV03	CCV	04/16/25 16:10		Kareem	OK
37	CCB03	CCB03	CCB	04/16/25 16:14		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135463

Review By	jaswal	Review On	4/18/2025 11:44:26 AM
Supervise By	Janvi	Supervise On	4/18/2025 12:06:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85016,MP85022,MP85020,MP85019,MP85018,MP85017
ICV Standard	MP85023,MP85022
CCV Standard	MP85026
ICSA Standard	MP85024,MP85025
CRI Standard	MP85022
LCS Standard	
Chk Standard	MP85030,MP85031

38	Q1803-01	WEST-BAY	SAM	04/16/25 16:19		Kareem	OK
39	Q1803-02	FUEL-MONITORING	SAM	04/16/25 16:23		Kareem	OK
40	Q1808-02	OILY-SOIL-PILE	SAM	04/16/25 16:28		Kareem	OK
41	Q1808-04	LAW-25-0060	SAM	04/16/25 16:32		Kareem	OK
42	Q1808-04DUP	LAW-25-0060DUP	DUP	04/16/25 16:36		Kareem	OK
43	Q1808-04L	LAW-25-0060L	SD	04/16/25 16:41		Kareem	OK
44	Q1808-04MS	LAW-25-0060MS	MS	04/16/25 16:45		Kareem	OK
45	Q1808-04MSD	LAW-25-0060MSD	MSD	04/16/25 16:49		Kareem	OK
46	Q1808-04A	LAW-25-0060A	PS	04/16/25 16:54	0.1ML EACH M6004,M6013-10ML SAMPLE	Kareem	OK
47	PB167587TB	PB167587TB	MB	04/16/25 16:58		Kareem	OK
48	CCV04	CCV04	CCV	04/16/25 17:02		Kareem	OK
49	CCB04	CCB04	CCB	04/16/25 17:06		Kareem	OK
50	PB167605BL	PB167605BL	MB	04/16/25 17:11		Kareem	OK
51	PB167605BS	PB167605BS	LCS	04/16/25 17:15		Kareem	OK
52	Q1791-09	VB16193	SAM	04/16/25 17:19		Kareem	OK
53	Q1791-10	280722	SAM	04/16/25 17:24		Kareem	OK
54	Q1791-11	280791	SAM	04/16/25 17:28		Kareem	OK
55	Q1791-14	VB16164	SAM	04/16/25 17:37		Kareem	OK
56	Q1795-02	31125	SAM	04/16/25 17:41	Not on login page	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135463

Review By	jaswal	Review On	4/18/2025 11:44:26 AM
Supervise By	Janvi	Supervise On	4/18/2025 12:06:47 PM

STD. NAME	STD REF.#
ICAL Standard	MP85016,MP85022,MP85020,MP85019,MP85018,MP85017
ICV Standard	MP85023,MP85022
CCV Standard	MP85026
ICSA Standard	MP85024,MP85025
CRI Standard	MP85022
LCS Standard	
Chk Standard	MP85030,MP85031

57	Q1795-03DL	10824-ADL	SAM	04/16/25 17:46	Not on login page	Kareem	Not Ok
58	Q1810-02	MOO-25-0118	SAM	04/16/25 17:50		Kareem	OK
59	CCV05	CCV05	CCV	04/16/25 17:55		Kareem	OK
60	CCB05	CCB05	CCB	04/16/25 17:59		Kareem	OK
61	Q1810-01DL	MOO-25-0117DL	SAM	04/16/25 18:03		Kareem	OK
62	Q1810-01DUPDL	MOO-25-0117DUPDL	DUP	04/16/25 18:08		Kareem	OK
63	Q1810-01LDL	MOO-25-0117LDL	SD	04/16/25 18:12		Kareem	OK
64	Q1810-01MSDL	MOO-25-0117MSDL	MS	04/16/25 18:16		Kareem	OK
65	Q1810-01MSDDL	MOO-25-0117MSDDL	MSD	04/16/25 18:21		Kareem	OK
66	Q1810-01ADL	MOO-25-0117ADL	PS	04/16/25 18:25		Kareem	OK
67	Q1791-12	V1207	SAM	04/16/25 18:30	Ag oversaturated	Kareem	Dilution
68	Q1810-01DL2	MOO-25-0117DL2	SAM	04/16/25 18:35	Not use	Kareem	Not Ok
69	Q1810-01DUPDL2	MOO-25-0117DUPDL	DUP	04/16/25 18:39	Not use	Kareem	Not Ok
70	Q1810-01LDL2	MOO-25-0117LDL2	SD	04/16/25 18:43	Not use	Kareem	Not Ok
71	CCV06	CCV06	CCV	04/16/25 18:48		Kareem	OK
72	CCB06	CCB06	CCB	04/16/25 18:52		Kareem	OK
73	Q1810-01MSDL2	MOO-25-0117MSDL2	MS	04/16/25 18:56	Not use	Kareem	Not Ok
74	Q1810-01MSDDL2	MOO-25-0117MSDDL	MSD	04/16/25 19:00	Not use	Kareem	Not Ok
75	Q1810-01ADL2	MOO-25-0117ADL2	PS	04/16/25 19:05	Not use	Kareem	Not Ok
76	PB167613BL	PB167613BL	MB	04/16/25 19:09		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135463

Review By	jaswal	Review On	4/18/2025 11:44:26 AM
Supervise By	Janvi	Supervise On	4/18/2025 12:06:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85016,MP85022,MP85020,MP85019,MP85018,MP85017		
ICV Standard	MP85023,MP85022		
CCV Standard	MP85026		
ICSA Standard	MP85024,MP85025		
CRI Standard	MP85022		
LCS Standard			
Chk Standard	MP85030,MP85031		

77	PB167613BS	PB167613BS	LCS	04/16/25 19:13		Kareem	OK
78	Q1791-12DL	V1207DL	SAM	04/16/25 19:17	5X for Ag	Kareem	Confirms
79	Q1804-01	001-WILLETS-PT-BL	SAM	04/16/25 19:22		Kareem	OK
80	Q1804-02	002-35TH-AVE(APR)	SAM	04/16/25 19:26		Kareem	OK
81	Q1804-02DUP	002-35TH-AVE(APR)	DUP	04/16/25 19:30		Kareem	OK
82	Q1804-02L	002-35TH-AVE(APR)	SD	04/16/25 19:35		Kareem	OK
83	CCV07	CCV07	CCV	04/16/25 19:39		Kareem	OK
84	CCB07	CCB07	CCB	04/16/25 19:43		Kareem	OK
85	Q1804-02MS	002-35TH-AVE(APR)	MS	04/16/25 19:48		Kareem	OK
86	Q1804-02MSD	002-35TH-AVE(APR)	MSD	04/16/25 19:52		Kareem	OK
87	Q1804-02A	002-35TH-AVE(APR)	PS	04/16/25 19:56		Kareem	OK
88	Q1815-01	001-WILLETS-PT-BL	SAM	04/16/25 20:00		Kareem	OK
89	Q1815-02	002-35TH-AVE(MAR)	SAM	04/16/25 20:04		Kareem	OK
90	CCV08	CCV08	CCV	04/16/25 20:08		Kareem	OK
91	CCB08	CCB08	CCB	04/16/25 20:13		Kareem	OK