



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Scalamandre – Tully JV **SDG No.:** Q1803
Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q1803 **SAS No.:** Q1803
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	98	40 - 160	CV	04/16/2025	13:22	LB135455
CRI	Arsenic	23.3	20.0	116	40 - 160	P	04/16/2025	13:46	LB135463
	Barium	91.3	100	91	40 - 160	P	04/16/2025	13:46	LB135463
	Cadmium	6.00	6.0	100	40 - 160	P	04/16/2025	13:46	LB135463
	Chromium	9.57	10.0	96	40 - 160	P	04/16/2025	13:46	LB135463
	Lead	12.9	12.0	107	40 - 160	P	04/16/2025	13:46	LB135463
	Selenium	19.0	20.0	95	40 - 160	P	04/16/2025	13:46	LB135463
	Silver	9.95	10.0	100	40 - 160	P	04/16/2025	13:46	LB135463