



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Scalamandre – Tully JV				SDG No.:	Q1803				
Contract:	SCAL01	Lab Code:	CHEM		Case No.:	Q1803		SAS No.:	Q1803	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB90	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:15	LB135455	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q1803

Contract: SCAL01

Lab Code: CHEM

Case No.: Q1803

SAS No.: Q1803

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:20	LB135455
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:52	LB135455
CCB81	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	14:20	LB135455

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q1803

Contract: SCAL01

Lab Code: CHEM

Case No.: Q1803

SAS No.: Q1803

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	13:42	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	13:42	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	13:42	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	13:42	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	13:42	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	13:42	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	13:42	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV		SDG No.: Q1803							
Contract: SCAL01	Lab Code: CHEM	Case No.: Q1803	SAS No.: Q1803						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	14:12	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	14:12	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	14:12	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	14:12	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	14:12	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	14:12	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	14:12	LB135463
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	15:23	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	15:23	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	15:23	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	15:23	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	15:23	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	15:23	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	15:23	LB135463
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	16:14	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	16:14	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	16:14	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	16:14	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	16:14	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	16:14	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	16:14	LB135463
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	17:06	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	17:06	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	17:06	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	17:06	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	17:06	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	17:06	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	17:06	LB135463
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	17:59	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	17:59	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	17:59	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	17:59	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	17:59	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	17:59	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	17:59	LB135463
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	18:52	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	18:52	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	18:52	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	18:52	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Scalamandre – Tully JV

SDG No.: Q1803

Contract: SCAL01

Lab Code: CHEM

Case No.: Q1803

SAS No.: Q1803

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	18:52	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	18:52	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	18:52	LB135463
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	19:43	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	19:43	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	19:43	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	19:43	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	19:43	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	19:43	LB135463
CCB08	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	19:43	LB135463
	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	20:13	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	20:13	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	20:13	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	20:13	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	20:13	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	20:13	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	20:13	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Scalamandre – Tully JV</u>				SDG No.:	<u>Q1803</u>			
Contract:	<u>SCAL01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1803</u>	SAS No.:	<u>Q1803</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number