

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Scalamandre – Tully JV</u>	SDG No.: <u>Q1803</u>
Contract: <u>SCAL01</u> Lab Code: <u>CHEM</u>	Case No.: <u>Q1803</u> SAS No.: <u>Q1803</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-14.1			-20	20	04/16/2025	13:51	LB135463
	Barium	1.88	6.0	31	-94	106	04/16/2025	13:51	LB135463
	Cadmium	-1.99	1.0	199	-5	7	04/16/2025	13:51	LB135463
	Chromium	55.8	52.0	107	42	62	04/16/2025	13:51	LB135463
	Lead	-4.75			-12	12	04/16/2025	13:51	LB135463
	Selenium	-3.36			-20	20	04/16/2025	13:51	LB135463
	Silver	-4.60			-10	10	04/16/2025	13:51	LB135463
ICSAB01	Arsenic	94.0	104	90	88.4	120	04/16/2025	13:55	LB135463
	Barium	468	537	87	437	637	04/16/2025	13:55	LB135463
	Cadmium	1040	972	107	826	1120	04/16/2025	13:55	LB135463
	Chromium	554	542	102	460	624	04/16/2025	13:55	LB135463
	Lead	43.2	49.0	88	37	61	04/16/2025	13:55	LB135463
	Selenium	47.0	46.0	102	26	66	04/16/2025	13:55	LB135463
	Silver	179	201	89	170	232	04/16/2025	13:55	LB135463