

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Scalamandre – Tully JV **Level:** LOW **SDG No.:** Q1803
Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q1803 **SAS No.:** Q1803
Matrix: Water **Sample ID:** Q1808-04 **Client ID:** LAW-25-0060DUP
Percent Solids for Sample: NA **Duplicate ID** Q1808-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	997		996		0		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	0.80	J	0.76	J	5		CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Scalamandre – Tully JV **Level:** LOW **SDG No.:** Q1803
Contract: SCAL01 **Lab Code:** CHEM **Case No.:** Q1803 **SAS No.:** Q1803
Matrix: Water **Sample ID:** Q1808-04MS **Client ID:** LAW-25-0060MSD
Percent Solids for Sample: NA **Duplicate ID** Q1808-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3880		3890		0		P
Barium	ug/L	20	1920		1890		2		P
Cadmium	ug/L	20	967		961		1		P
Chromium	ug/L	20	1910		1910		0		P
Lead	ug/L	20	4620		4580		1		P
Mercury	ug/L	20	39.2		40.1		2		CV
Selenium	ug/L	20	9470		9430		0		P
Silver	ug/L	20	349		347		1		P