

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: SCAL01
Lab Code: CHEM **Case No.:** Q1803 **SAS No.:** Q1803 **SDG NO.:** Q1803
Instrument ID: ECD_S **Calibration Date(s):** 04/02/2025 04/02/2025
Calibration Times: 17:32 20:44

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029657.D</u>	CF 500 =	<u>PS029658.D</u>		
CF 750 =		<u>PS029659.D</u>	CF 1000 =	<u>PS029660.D</u>	CF 1500 =	<u>PS029661.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	12727900000	11569300000	11320400000	11249800000	11209500000	11615400000	5
2,4-D	2821030000	2334770000	2267580000	2246770000	2377810000	2409590000	10
2,4-DCAA	2302820000	2069000000	1979340000	1949030000	1893970000	2038830000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

Instrument ID: ECD_S

Calibration Date(s): 04/02/2025 04/02/2025

Calibration Times: 17:32 20:44

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029657.D</u>	CF 500 =	<u>PS029658.D</u>		
CF 750 =		<u>PS029659.D</u>	CF 1000 =	<u>PS029660.D</u>	CF 1500 =	<u>PS029661.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	7615620000	7063890000	7056380000	7090810000	7238100000	7212960000	3
2,4-D	1263680000	1033640000	1003030000	1003730000	1088250000	1078460000	10
2,4-DCAA	756673000	677593000	660102000	658637000	654559000	681513000	6