

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: HSTDCCC750 Data File : PS029836.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.872	8.788	8.988	741.060	712.500	4.0
2,4-D	8.028	7.942	8.142	701.300	705.000	-0.5
2,4-DCAA	6.950	6.863	7.063	751.320	750.000	0.2

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: HSTDCCC750 Data File : PS029836.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.550	9.451	9.651	740.970	712.500	4.0
2,4-D	8.669	8.571	8.771	668.580	705.000	-5.2
2,4-DCAA	7.474	7.375	7.575	705.620	750.000	-5.9