

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: HSTDCCC750 Data File : PS029839.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	8.872	8.788	8.988	737.760	712.500	3.5
2,4-D	8.028	7.942	8.142	697.010	705.000	-1.1
2,4-DCAA	6.950	6.863	7.063	745.960	750.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: HSTDCCC750 Data File : PS029839.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.548	9.451	9.651	734.150	712.500	3.0
2,4-D	8.668	8.571	8.771	662.800	705.000	-6.0
2,4-DCAA	7.473	7.375	7.575	705.980	750.000	-5.9