

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040225

Review By	Abdul	Review On	4/3/2025 9:25:38 AM
Supervise By	mohammad	Supervise On	4/4/2025 7:53:55 AM
SubDirectory	PS040225	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029655.D	02 Apr 2025 16:20		AR\AJ	Ok
2	I.BLK	I.BLK	PS029656.D	02 Apr 2025 16:44		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029657.D	02 Apr 2025 17:32		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029658.D	02 Apr 2025 17:56		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029659.D	02 Apr 2025 18:44		AR\AJ	Ok,M
6	HSTDICC1000	HSTDICC1000	PS029660.D	02 Apr 2025 19:32		AR\AJ	Ok,M
7	HSTDICC1500	HSTDICC1500	PS029661.D	02 Apr 2025 20:44		AR\AJ	Ok,M
8	HSTDICV750	ICVPS040225	PS029662.D	02 Apr 2025 21:32		AR\AJ	Ok,M
9	I.BLK	I.BLK	PS029663.D	02 Apr 2025 21:56		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS041625

Review By	Abdul	Review On	4/17/2025 8:05:57 AM
Supervise By	mohammad	Supervise On	4/18/2025 1:19:34 AM
SubDirectory	PS041625	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029821.D	16 Apr 2025 14:41		AR\AJ	Ok
2	I.BLK	I.BLK	PS029822.D	16 Apr 2025 15:05		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS029823.D	16 Apr 2025 16:48		AR\AJ	Ok,M
4	PB167608BL	PB167608BL	PS029824.D	16 Apr 2025 17:12		AR\AJ	Ok,M
5	PB167608BS	PB167608BS	PS029825.D	16 Apr 2025 17:36	Comp#2,8,9,10,11,12,16 recovery fail	AR\AJ	Not Ok
6	PB167587TB	PB167587TB	PS029826.D	16 Apr 2025 18:01		AR\AJ	Ok,M
7	Q1803-01	WEST-BAY	PS029827.D	16 Apr 2025 18:25		AR\AJ	Ok,M
8	Q1803-02	FUEL-MONITORING	PS029828.D	16 Apr 2025 18:49		AR\AJ	Ok
9	Q1800-03	WC-A4-01-C	PS029829.D	16 Apr 2025 19:13		AR\AJ	Ok,M
10	Q1800-03MS	WC-A4-01-CMS	PS029830.D	16 Apr 2025 19:37	some compound recovery fail	AR\AJ	Ok,M
11	Q1800-03MSD	WC-A4-01-CMSD	PS029831.D	16 Apr 2025 20:01	some compound recovery fail	AR\AJ	Ok,M
12	I.BLK	I.BLK	PS029832.D	16 Apr 2025 20:25		AR\AJ	Ok
13	HSTDCCC750	HSTDCCC750	PS029833.D	16 Apr 2025 20:49		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS041725

Review By	Abdul	Review On	4/18/2025 7:42:09 AM
Supervise By		Supervise On	
SubDirectory	PS041725	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029834.D	17 Apr 2025 08:13		AR\AJ	Ok
2	I.BLK	I.BLK	PS029835.D	17 Apr 2025 08:37		AR\AJ	Ok,NS
3	HSTDCCC750	HSTDCCC750	PS029836.D	17 Apr 2025 09:01		AR\AJ	Ok,NS
4	PB167608BS	PB167608BS	PS029837.D	17 Apr 2025 12:55		AR\AJ	Ok,NS
5	I.BLK	I.BLK	PS029838.D	17 Apr 2025 15:37		AR\AJ	Ok
6	HSTDCCC750	HSTDCCC750	PS029839.D	17 Apr 2025 16:01		AR\AJ	Ok,NS
7	PB167620BL	PB167620BL	PS029840.D	17 Apr 2025 18:39		AR\AJ	Ok,NS
8	PB167620BS	PB167620BS	PS029841.D	17 Apr 2025 19:03	DCCA Recovery high	AR\AJ	Not Ok
9	Q1787-09	TP-1	PS029842.D	17 Apr 2025 19:27		AR\AJ	Ok,NS
10	Q1788-01	WC-1	PS029843.D	17 Apr 2025 19:51		AR\AJ	Ok,NS
11	Q1788-01MS	WC-1MS	PS029844.D	17 Apr 2025 20:15	F Flag coming , some compound recovery fail	AR\AJ	Not Ok
12	Q1788-01MSD	WC-1MSD	PS029845.D	17 Apr 2025 20:39	F Flag coming , some compound recovery fail , RPD Fail	AR\AJ	Not Ok
13	Q1808-01	OILY-SOIL-PILE	PS029846.D	17 Apr 2025 21:03		AR\AJ	Ok
14	Q1808-03	LAW-25-0060	PS029847.D	17 Apr 2025 21:27		AR\AJ	Ok,NS
15	I.BLK	I.BLK	PS029848.D	17 Apr 2025 21:51		AR\AJ	Ok
16	HSTDCCC750	HSTDCCC750	PS029849.D	17 Apr 2025 22:15		AR\AJ	Ok,NS

M : Manual Integration