

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL01 Date Analyzed: 04/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095256.D Time Analyzed: 11:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.064	8.953	9.153	57.350	50.000	14.7
Endrin	6.579	6.471	6.671	53.700	50.000	7.4
gamma-BHC (Lindane)	4.332	4.223	4.423	48.540	50.000	-2.9
Heptachlor	4.920	4.811	5.011	50.510	50.000	1.0
Heptachlor epoxide	5.689	5.580	5.780	51.780	50.000	3.6
Methoxychlor	7.508	7.398	7.598	49.260	50.000	-1.5
Tetrachloro-m-xylene	3.542	3.435	3.635	57.180	50.000	14.4

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL01 Date Analyzed: 04/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095256.D Time Analyzed: 11:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.903	7.800	8.000	58.920	50.000	17.8
Endrin	5.629	5.527	5.727	54.020	50.000	8.0
gamma-BHC (Lindane)	3.599	3.499	3.699	47.120	50.000	-5.8
Heptachlor	3.937	3.836	4.036	49.420	50.000	-1.2
Heptachlor epoxide	4.719	4.618	4.818	50.540	50.000	1.1
Methoxychlor	6.603	6.499	6.699	51.320	50.000	2.6
Tetrachloro-m-xylene	2.767	2.667	2.867	55.800	50.000	11.6