

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL095270.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.050	8.953	9.153	56.810	50.000	13.6
Endrin	6.569	6.471	6.671	53.280	50.000	6.6
gamma-BHC (Lindane)	4.326	4.223	4.423	49.090	50.000	-1.8
Heptachlor	4.913	4.811	5.011	50.500	50.000	1.0
Heptachlor epoxide	5.681	5.580	5.780	51.800	50.000	3.6
Methoxychlor	7.497	7.398	7.598	47.650	50.000	-4.7
Tetrachloro-m-xylene	3.536	3.435	3.635	57.770	50.000	15.5

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL03 Date Analyzed: 04/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL095270.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.898	7.800	8.000	58.790	50.000	17.6
Endrin	5.628	5.527	5.727	53.110	50.000	6.2
gamma-BHC (Lindane)	3.600	3.499	3.699	47.870	50.000	-4.3
Heptachlor	3.937	3.836	4.036	50.220	50.000	0.4
Heptachlor epoxide	4.719	4.618	4.818	50.880	50.000	1.8
Methoxychlor	6.600	6.499	6.699	49.610	50.000	-0.8
Tetrachloro-m-xylene	2.769	2.667	2.867	56.240	50.000	12.5