

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL095273.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.049	8.953	9.153	55.330	50.000	10.7
Endrin	6.569	6.471	6.671	51.700	50.000	3.4
gamma-BHC (Lindane)	4.324	4.223	4.423	47.530	50.000	-4.9
Heptachlor	4.911	4.811	5.011	49.100	50.000	-1.8
Heptachlor epoxide	5.680	5.580	5.780	50.450	50.000	0.9
Methoxychlor	7.497	7.398	7.598	46.880	50.000	-6.2
Tetrachloro-m-xylene	3.536	3.435	3.635	56.370	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Contract: SCAL01

Lab Code: CHEM Case No.: Q1803 SAS No.: Q1803 SDG NO.: Q1803

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL04 Date Analyzed: 04/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL095273.D Time Analyzed: 10:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.898	7.800	8.000	57.830	50.000	15.7
Endrin	5.626	5.527	5.727	51.830	50.000	3.7
gamma-BHC (Lindane)	3.599	3.499	3.699	46.520	50.000	-7.0
Heptachlor	3.936	3.836	4.036	48.880	50.000	-2.2
Heptachlor epoxide	4.717	4.618	4.818	49.620	50.000	-0.8
Methoxychlor	6.599	6.499	6.699	48.820	50.000	-2.4
Tetrachloro-m-xylene	2.768	2.667	2.867	54.510	50.000	9.0