

Surrogate Summary

SDG No.: **Q1803**

Client: **Scalamandre – Tully JV**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095202.D	PIBLK-PL095202.D	Decachlorobiphenyl	1	20	22.6	113		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.7	109		43	140
		Tetrachloro-m-xylene	2	20	20.1	100		77	126
I.BLK-PL095254.D	PIBLK-PL095254.D	Decachlorobiphenyl	1	20	21.1	106		43	140
		Tetrachloro-m-xylene	1	20	20.9	104		77	126
		Decachlorobiphenyl	2	20	19.4	97		43	140
		Tetrachloro-m-xylene	2	20	19.3	97		77	126
PB167609BL	PB167609BL	Decachlorobiphenyl	1	20	20.6	103		43	140
		Tetrachloro-m-xylene	1	20	18.9	94		77	126
		Decachlorobiphenyl	2	20	20.0	100		43	140
		Tetrachloro-m-xylene	2	20	17.4	87		77	126
PB167609BS	PB167609BS	Decachlorobiphenyl	1	20	20.1	100		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	19.4	97		43	140
		Tetrachloro-m-xylene	2	20	17.4	87		77	126
PB167587TB	PB167587TB	Decachlorobiphenyl	1	20	20.7	104		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	20.5	103		43	140
		Tetrachloro-m-xylene	2	20	18.5	92		77	126
Q1803-01	WEST-BAY	Decachlorobiphenyl	1	20	20.1	100		43	140
		Tetrachloro-m-xylene	1	20	14.1	71	*	77	126
		Decachlorobiphenyl	2	20	20.1	100		43	140
		Tetrachloro-m-xylene	2	20	12.9	65	*	77	126
Q1803-02	FUEL-MONITORING	Decachlorobiphenyl	1	20	21.0	105		43	140
		Tetrachloro-m-xylene	1	20	17.5	88		77	126
		Decachlorobiphenyl	2	20	20.8	104		43	140
		Tetrachloro-m-xylene	2	20	15.9	80		77	126
Q1800-03MS	WC-A4-01-CMS	Decachlorobiphenyl	1	20	22.4	112		43	140
		Tetrachloro-m-xylene	1	20	19.5	98		77	126
		Decachlorobiphenyl	2	20	22.5	113		43	140
		Tetrachloro-m-xylene	2	20	16.8	84		77	126
Q1800-03MSD	WC-A4-01-CMSD	Decachlorobiphenyl	1	20	22.2	111		43	140
		Tetrachloro-m-xylene	1	20	19.3	97		77	126
		Decachlorobiphenyl	2	20	22.4	112		43	140
		Tetrachloro-m-xylene	2	20	16.8	84		77	126
I.BLK-PL095265.D	PIBLK-PL095265.D	Decachlorobiphenyl	1	20	21.0	105		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	20.6	103		43	140
		Tetrachloro-m-xylene	2	20	19.1	96		77	126
I.BLK-PL095268.D	PIBLK-PL095268.D	Decachlorobiphenyl	1	20	21.2	106		43	140

Surrogate Summary

SDG No.: Q1803

Client: Scalamandre – Tully JV

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL095268.D	PIBLK-PL095268.D	Tetrachloro-m-xylene	1	20	21.2	106		77	126
		Decachlorobiphenyl	2	20	20.1	101		43	140
Q1803-01RE	WEST-BAYRE	Tetrachloro-m-xylene	2	20	19.7	99		77	126
		Decachlorobiphenyl	1	20	20.3	102		43	140
		Tetrachloro-m-xylene	1	20	14.4	72	*	77	126
		Decachlorobiphenyl	2	20	20.0	100		43	140
		Tetrachloro-m-xylene	2	20	12.9	65	*	77	126
I.BLK-PL095272.D	PIBLK-PL095272.D	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	20.4	102		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126