

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM040825

Review By	anahy	Review On	4/9/2025 10:38:14 AM		
Supervise By	Jagrut	Supervise On	4/9/2025 4:50:06 PM		
SubDirectory	BM040825	HP Acquire Method	BNA_M	HP Processing Method	bm040825
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12657,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM049847.D	08 Apr 2025 12:55	RC/JU	Ok
2	SSTDICC2.5	BM049848.D	08 Apr 2025 13:35	RC/JU	Ok
3	SSTDICC005	BM049849.D	08 Apr 2025 14:14	RC/JU	Ok,M
4	SSTDICC010	BM049850.D	08 Apr 2025 14:53	RC/JU	Ok,M
5	SSTDICC020	BM049851.D	08 Apr 2025 15:32	RC/JU	Ok
6	SSTDICCC040	BM049852.D	08 Apr 2025 16:12	RC/JU	Ok
7	SSTDICC050	BM049853.D	08 Apr 2025 17:30	RC/JU	Ok
8	SSTDICC060	BM049854.D	08 Apr 2025 19:28	RC/JU	Ok,M
9	SSTDICC080	BM049855.D	08 Apr 2025 20:07	RC/JU	Ok,M
10	SSTDICV040	BM049856.D	08 Apr 2025 20:47	RC/JU	Ok
11	PB167474BL	BM049857.D	08 Apr 2025 21:26	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM041625

Review By	Rahul	Review On	4/17/2025 10:26:53 AM		
Supervise By	Jagrut	Supervise On	4/17/2025 12:28:46 PM		
SubDirectory	BM041625	HP Acquire Method	BNA_M	HP Processing Method	bm040825
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12659,10ul/1000ul sample			
ICV/I.BLK		SP6686			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM049943.D	16 Apr 2025 08:49	RC/JU	Ok
2	SSTDCCC040	BM049944.D	16 Apr 2025 09:28	RC/JU	Ok,M
3	PB167561BL	BM049945.D	16 Apr 2025 10:07	RC/JU	Ok
4	PB167561BS	BM049946.D	16 Apr 2025 10:47	RC/JU	Ok,M
5	PB167509BS	BM049947.D	16 Apr 2025 11:26	RC/JU	Ok,M
6	PB167587TB	BM049948.D	16 Apr 2025 12:05	RC/JU	Ok
7	PB167598BL	BM049949.D	16 Apr 2025 12:44	RC/JU	Ok
8	PB167598BS	BM049950.D	16 Apr 2025 13:24	RC/JU	Ok,M
9	Q1791-13	BM049951.D	16 Apr 2025 14:07	RC/JU	ReRun
10	Q1791-14	BM049952.D	16 Apr 2025 14:46	RC/JU	Ok
11	Q1810-02	BM049953.D	16 Apr 2025 15:26	RC/JU	Ok
12	Q1806-01DL	BM049954.D	16 Apr 2025 16:05	RC/JU	Ok,M
13	Q1788-01DL	BM049955.D	16 Apr 2025 16:44	RC/JU	Ok,M
14	Q1808-01MS	BM049956.D	16 Apr 2025 17:24	RC/JU	Ok,M
15	Q1808-01MSD	BM049957.D	16 Apr 2025 18:03	RC/JU	Ok,M
16	Q1808-03	BM049958.D	16 Apr 2025 18:42	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM		
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM		
SubDirectory	BP041425	HP Acquire Method	BNA_P	HP Processing Method	bp041425
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12659,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024274.D	14 Apr 2025 10:25	RC/JU	Ok
2	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06	RC/JU	Ok
3	SSTDICC005	BP024276.D	14 Apr 2025 11:47	RC/JU	Ok,M
4	SSTDICC010	BP024277.D	14 Apr 2025 12:27	RC/JU	Ok,M
5	SSTDICC020	BP024278.D	14 Apr 2025 13:08	RC/JU	Ok,M
6	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	RC/JU	Ok
7	SSTDICC050	BP024280.D	14 Apr 2025 15:10	RC/JU	Ok,M
8	SSTDICC060	BP024281.D	14 Apr 2025 16:32	RC/JU	Ok,M
9	SSTDICC080	BP024282.D	14 Apr 2025 17:13	RC/JU	Ok,M
10	SSTDICV040	BP024283.D	14 Apr 2025 18:35	RC/JU	Ok,M
11	PB167488TB	BP024284.D	14 Apr 2025 19:16	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP041525

Review By	Rahul	Review On	4/16/2025 9:40:54 AM
Supervise By	Jagrut	Supervise On	4/17/2025 3:00:43 PM
SubDirectory	BP041525	HP Acquire Method	BNA_P
		HP Processing Method	bp041425
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12659,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024285.D	15 Apr 2025 08:34	RC/JU	Ok
2	SSTDCCC040	BP024286.D	15 Apr 2025 09:15	RC/JU	Ok,M
3	PB167571BL	BP024287.D	15 Apr 2025 09:56	RC/JU	Ok
4	PB167571BS	BP024288.D	15 Apr 2025 10:36	RC/JU	Ok,M
5	PB167547TB	BP024289.D	15 Apr 2025 11:17	RC/JU	Ok
6	Q1771-04	BP024290.D	15 Apr 2025 12:02	RC/JU	Ok
7	Q1771-08	BP024291.D	15 Apr 2025 12:43	RC/JU	Ok
8	Q1771-12	BP024292.D	15 Apr 2025 13:24	RC/JU	Ok
9	Q1779-04	BP024293.D	15 Apr 2025 14:05	RC/JU	Ok
10	Q1779-08	BP024294.D	15 Apr 2025 14:46	RC/JU	Ok
11	Q1781-04	BP024295.D	15 Apr 2025 15:27	RC/JU	Ok
12	Q1781-08	BP024296.D	15 Apr 2025 16:08	RC/JU	Ok
13	Q1781-12	BP024297.D	15 Apr 2025 16:48	RC/JU	Ok
14	Q1787-12	BP024298.D	15 Apr 2025 17:29	RC/JU	Ok
15	Q1787-12MS	BP024299.D	15 Apr 2025 18:10	RC/JU	Ok,M
16	Q1787-12MSD	BP024300.D	15 Apr 2025 18:51	RC/JU	Ok
17	Q1803-01	BP024301.D	15 Apr 2025 19:32	RC/JU	Ok
18	Q1803-02	BP024302.D	15 Apr 2025 20:13	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP041625

Review By	Rahul	Review On	4/17/2025 10:27:49 AM
Supervise By	Jagrut	Supervise On	4/17/2025 12:29:39 PM
SubDirectory	BP041625	HP Acquire Method	BNA_P
		HP Processing Method	bp041425
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12659,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP024303.D	16 Apr 2025 09:02	RC/JU	Ok
2	SSTDCCC040	BP024304.D	16 Apr 2025 09:43	RC/JU	Ok
3	PB167606BL	BP024305.D	16 Apr 2025 10:23	RC/JU	Ok
4	PB167606BS	BP024306.D	16 Apr 2025 11:04	RC/JU	Ok,M
5	PB167599BS	BP024307.D	16 Apr 2025 11:45	RC/JU	Ok
6	PB167599BSD	BP024308.D	16 Apr 2025 12:26	RC/JU	Ok,M
7	PB167564BL	BP024309.D	16 Apr 2025 13:07	RC/JU	Ok
8	Q1788-04	BP024310.D	16 Apr 2025 13:52	RC/JU	Ok
9	Q1800-03	BP024311.D	16 Apr 2025 14:32	RC/JU	Ok
10	Q1808-02	BP024312.D	16 Apr 2025 15:13	RC/JU	Ok
11	Q1808-04	BP024313.D	16 Apr 2025 15:54	RC/JU	Ok
12	Q1791-09	BP024314.D	16 Apr 2025 16:35	RC/JU	Ok,M
13	Q1791-10	BP024315.D	16 Apr 2025 17:16	RC/JU	ReRun
14	Q1791-11	BP024316.D	16 Apr 2025 17:57	RC/JU	Ok
15	Q1791-12	BP024317.D	16 Apr 2025 18:38	RC/JU	Ok
16	Q1762-02	BP024318.D	16 Apr 2025 19:18	RC/JU	Ok,M

M : Manual Integration