

Surrogate Summary

SW-846

SDG No.: Q1803

Client: Scalamandre – Tully JV

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167587TB	PB167587TB	2-Fluorophenol	150	129	86		10	139
		Phenol-d6	150	124	83		10	134
		Nitrobenzene-d5	100	91.8	92		49	133
		2-Fluorobiphenyl	100	91.0	91		52	132
		2,4,6-Tribromophenol	150	142	95		44	137
		Terphenyl-d14	100	112	112		48	125
PB167606BL	PB167606BL	2-Fluorophenol	150	133	89		10	139
		Phenol-d6	150	119	80		10	134
		Nitrobenzene-d5	100	89.0	89		49	133
		2-Fluorobiphenyl	100	90.1	90		52	132
		2,4,6-Tribromophenol	150	118	79		44	137
		Terphenyl-d14	100	86.7	87		48	125
PB167606BS	PB167606BS	2-Fluorophenol	150	135	90		10	139
		Phenol-d6	150	123	82		10	134
		Nitrobenzene-d5	100	85.2	85		49	133
		2-Fluorobiphenyl	100	85.9	86		52	132
		2,4,6-Tribromophenol	150	124	83		44	137
		Terphenyl-d14	100	90.2	90		48	125
Q1787-12MS	TP-1MS	2-Fluorophenol	150	124	83		10	139
		Phenol-d6	150	103	68		10	134
		Nitrobenzene-d5	100	95.4	95		49	133
		2-Fluorobiphenyl	100	88.7	89		52	132
		2,4,6-Tribromophenol	150	125	83		44	137
		Terphenyl-d14	100	89.1	89		48	125
Q1787-12MSD	TP-1MSD	2-Fluorophenol	150	130	87		10	139
		Phenol-d6	150	109	72		10	134
		Nitrobenzene-d5	100	96.9	97		49	133
		2-Fluorobiphenyl	100	90.8	91		52	132
		2,4,6-Tribromophenol	150	133	88		44	137
		Terphenyl-d14	100	96.6	97		48	125
Q1803-01	WEST-BAY	2-Fluorophenol	150	115	76		10	139
		Phenol-d6	150	92.1	61		10	134
		Nitrobenzene-d5	100	92.2	92		49	133
		2-Fluorobiphenyl	100	83.9	84		52	132
		2,4,6-Tribromophenol	150	129	86		44	137
		Terphenyl-d14	100	96.8	97		48	125
Q1803-02	FUEL-MONITORING	2-Fluorophenol	150	123	82		10	139
		Phenol-d6	150	100	67		10	134
		Nitrobenzene-d5	100	96.0	96		49	133
		2-Fluorobiphenyl	100	86.0	86		52	132
		2,4,6-Tribromophenol	150	137	91		44	137
		Terphenyl-d14	100	101	101		48	125