

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: SCAL01

Lab Code: CHEM

SAS No.: Q1803

SDG NO.: Q1803

Lab File ID: BM049847.D

DFTPP Injection Date: 04/08/2025

Instrument ID: BNA_M

DFTPP Injection Time: 12:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.1
68	Less than 2.0% of mass 69	0.3 (1.2) 1
69	Mass 69 relative abundance	26.5
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	34.9
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.5
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	12.4
442	Greater than 50% of mass 198	81.2
443	15.0 - 24.0% of mass 442	15.3 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM049848.D	04/08/2025	13:35
SSTDICC005	SSTDICC005	BM049849.D	04/08/2025	14:14
SSTDICC010	SSTDICC010	BM049850.D	04/08/2025	14:53
SSTDICC020	SSTDICC020	BM049851.D	04/08/2025	15:32
SSTDICCC040	SSTDICCC040	BM049852.D	04/08/2025	16:12
SSTDICC050	SSTDICC050	BM049853.D	04/08/2025	17:30
SSTDICC060	SSTDICC060	BM049854.D	04/08/2025	19:28
SSTDICC080	SSTDICC080	BM049855.D	04/08/2025	20:07