

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1803

Client: Scalamandre – Tully JV

Analytical Method: SW8260-Low

Datafile : VX045800.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0416WBS01	Vinyl chloride	20	17.8	ug/L	89			65	117	
	1,1-Dichloroethene	20	18.4	ug/L	92			74	110	
	2-Butanone	100	98.8	ug/L	99			65	122	
	Carbon Tetrachloride	20	19.9	ug/L	100			77	113	
	Chloroform	20	19.7	ug/L	99			79	113	
	Benzene	20	18.9	ug/L	95			82	109	
	1,2-Dichloroethane	20	20.3	ug/L	102			80	115	
	Trichloroethene	20	19.3	ug/L	97			77	113	
	Tetrachloroethene	20	19.2	ug/L	96			67	123	
	Chlorobenzene	20	19.4	ug/L	97			82	109	