

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1804
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1804 **SAS No.:** Q1804
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	522	510	102	95 - 105	P	04/16/2025	13:09	LB135463
	Iron	9740	10000	97	95 - 105	P	04/16/2025	13:09	LB135463
	Lead	1010	1000	101	95 - 105	P	04/16/2025	13:09	LB135463

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1804
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1804 **SAS No.:** Q1804
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV1	Copper	22.1	20.0	110	80 - 120	P	04/16/2025	13:37	LB135463
	Iron	95.0	100	95	80 - 120	P	04/16/2025	13:37	LB135463
	Lead	11.7	12.0	97	80 - 120	P	04/16/2025	13:37	LB135463

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1804
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1804 **SAS No.:** Q1804
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	04/16/2025	14:08	LB135463
	Iron	4860	5000	97	90 - 110	P	04/16/2025	14:08	LB135463
	Lead	4920	5000	98	90 - 110	P	04/16/2025	14:08	LB135463
CCV02	Copper	1250	1250	100	90 - 110	P	04/16/2025	15:18	LB135463
	Iron	4900	5000	98	90 - 110	P	04/16/2025	15:18	LB135463
	Lead	4940	5000	99	90 - 110	P	04/16/2025	15:18	LB135463
CCV03	Copper	1230	1250	98	90 - 110	P	04/16/2025	16:10	LB135463
	Iron	4830	5000	97	90 - 110	P	04/16/2025	16:10	LB135463
	Lead	4860	5000	97	90 - 110	P	04/16/2025	16:10	LB135463
CCV04	Copper	1220	1250	98	90 - 110	P	04/16/2025	17:02	LB135463
	Iron	4870	5000	97	90 - 110	P	04/16/2025	17:02	LB135463
	Lead	4860	5000	97	90 - 110	P	04/16/2025	17:02	LB135463
CCV05	Copper	1200	1250	96	90 - 110	P	04/16/2025	17:55	LB135463
	Iron	4680	5000	94	90 - 110	P	04/16/2025	17:55	LB135463
	Lead	4720	5000	94	90 - 110	P	04/16/2025	17:55	LB135463
CCV06	Copper	1200	1250	96	90 - 110	P	04/16/2025	18:48	LB135463
	Iron	4620	5000	92	90 - 110	P	04/16/2025	18:48	LB135463
	Lead	4740	5000	95	90 - 110	P	04/16/2025	18:48	LB135463
CCV07	Copper	1240	1250	99	90 - 110	P	04/16/2025	19:39	LB135463
	Iron	4800	5000	96	90 - 110	P	04/16/2025	19:39	LB135463
	Lead	4910	5000	98	90 - 110	P	04/16/2025	19:39	LB135463
CCV08	Copper	1210	1250	97	90 - 110	P	04/16/2025	20:08	LB135463
	Iron	4810	5000	96	90 - 110	P	04/16/2025	20:08	LB135463
	Lead	4820	5000	96	90 - 110	P	04/16/2025	20:08	LB135463