



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q1804
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1804 **SAS No.:** Q1804
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Copper	22.4	20.0	112	40 - 160	P	04/16/2025	13:46	LB135463
	Iron	93.3	100	93	40 - 160	P	04/16/2025	13:46	LB135463
	Lead	12.9	12.0	107	40 - 160	P	04/16/2025	13:46	LB135463