

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1804 SAS No.: Q1804 SDG No.: Q1804

Instrument ID: MSVOA_N Calibration Date/Time: 04/14/2025 09:48

Lab File ID: VN086257.D Init. Calib. Date(s): 04/11/2025 04/11/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:04 12:40

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.710	1.605	0.5	-6.14	
Toluene	2.012	1.915	0.4	-4.82	
Ethyl Benzene	2.232	2.098	0.1	-6	
m/p-Xylenes	0.830	0.780	0.3	-6.02	
o-Xylene	0.816	0.787	0.3	-3.55	
1,2-Dichloroethane-d4	2.926	2.921	0.01	-0.17	
Toluene-d8	1.694	1.689	0.01	-0.29	
4-Bromofluorobenzene	0.597	0.586	0.2	-1.84	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.