

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: Q1804 SAS No.: Q1804 SDG NO.: Q1804
Lab File ID: VN086257.D Date Analyzed: 04/14/2025
Instrument ID: MSVOA_N Time Analyzed: 09:48
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	32024	7.81	197357	9.09	173474	11.87
UPPER LIMIT	64048	8.312	394714	9.594	346948	12.365
LOWER LIMIT	16012	7.312	98678.5	8.594	86737	11.365
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (APR)	32875	7.81	201433	9.10	178259	11.87
002-35TH-AVE (APR)	26217	7.81	164361	9.10	144392	11.87
VN0414WBL01	34059	7.81	208636	9.10	183354	11.87
VN0414WBS01	36879	7.81	225158	9.10	200399	11.87
VN0414WBSD01	36062	7.82	212146	9.10	188770	11.87

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.