

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01  
 Lab Code: CHEM Case No.: Q1804 SAS No.: Q1804 SDG NO.: Q1804  
 Lab File ID: VN086257.D Date Analyzed: 04/14/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:48  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                           | IS4<br>AREA # | RT # |  |  |  |  |
|---------------------------|---------------|------|--|--|--|--|
| 12 HOUR STD               | 0             | 0    |  |  |  |  |
| UPPER LIMIT               | 0             |      |  |  |  |  |
| LOWER LIMIT               | 0             |      |  |  |  |  |
| EPA SAMPLE NO.            |               |      |  |  |  |  |
| 001-WILLETS-PT-BLVD (APR) | 0             | 0.00 |  |  |  |  |
| 002-35TH-AVE (APR)        | 0             | 0.00 |  |  |  |  |
| VN0414WBL01               | 0             | 0.00 |  |  |  |  |
| VN0414WBS01               | 0             | 0.00 |  |  |  |  |
| VN0414WBSD01              | 0             | 0.00 |  |  |  |  |

IS4 =

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.