



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

JACOBS Engineering Group, Inc.

Project Name: Former Schlumberger STC PTC Site D3868221

Project # N/A

Chemtech Project # Q1812

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

3 Water samples were received on 04/15/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Diesel Range Organics, Gasoline Range Organics, Mercury, Metals ICP-RCRA, Metals ICP-TAL, METALS RCRA, METALS-TAL, SVOC-TCL BNA -20, TPH GC and VOC-TCLVOA-10. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for RINSE-EB-TANK-041525 [Phenol-d6 - 14%]. This compound did not meet the NJDKQP criteria but met the in-house criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike for {PB167625BS} with File ID: BP024339.D met requirements for all samples except for 3,3-Dichlorobenzidine[63%], 4-Chloroaniline[31%]-These compounds did not meet the NJDKQP criteria but met the in-house criteria and For Hexachlorocyclopentadiene[190%]-This compound did not meet the NJDKQP criteria and in-house criteria. But associated samples have not positive hit for these compounds therefore no corrective action was taken.

The Blank Spike Duplicate for {PB167625BSD} with File ID: BP024340.D met requirements for all samples except for 2,4-Dimethylphenol[132%], 3,3-Dichlorobenzidine[65%], 4-Chloroaniline[34%]-These compounds did not meet the NJDKQP criteria but met the in-house criteria and for Hexachlorocyclopentadiene[190%]-This compound did not meet the NJDKQP criteria



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

and in-house criteria But associated samples have not positive hit for these compounds therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID BP024320.D met the requirements except for Benzaldehyde . But associated samples have not positive hit for this compound therefore no corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____