



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB92	Mercury	0.20	+/-0.20	U	0.20	CV	04/17/2025	12:17	LB135470

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB85	Mercury	0.20	+/-0.20	U	0.20	CV	04/17/2025	12:22	LB135470
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	04/17/2025	12:55	LB135470
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	04/17/2025	13:25	LB135470

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	2.95	+/-20.0	J	20.0	P	04/21/2025	16:01	LB135513
	Antimony	2.00	+/-2.00	U	2.00	P	04/21/2025	16:01	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	16:01	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	16:01	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	16:01	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	16:01	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	16:01	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	16:01	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	16:01	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	16:01	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	16:01	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	16:01	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	16:01	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	16:01	LB135513

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	19.7	+/-20.0	J	20.0	P	04/21/2025	16:18	LB135513
	Antimony	2.00	+/-2.00	U	2.00	P	04/21/2025	16:18	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	16:18	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	16:18	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	16:18	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	16:18	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	16:18	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	16:18	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	16:18	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	16:18	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	16:18	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	16:18	LB135513
	Thallium	0.070	+/-1.00	J	1.00	P	04/21/2025	16:18	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	16:18	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	16:18	LB135513
CCB02	Aluminum	19.3	+/-20.0	J	20.0	P	04/21/2025	17:04	LB135513
	Antimony	2.00	+/-2.00	U	2.00	P	04/21/2025	17:04	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	17:04	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	17:04	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	17:04	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	17:04	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	17:04	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	17:04	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	17:04	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	17:04	LB135513

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	17:04	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:04	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	17:04	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	17:04	LB135513
CCB03	Aluminum	7.32	+/-20.0	J	20.0	P	04/21/2025	17:48	LB135513
	Antimony	2.00	+/-2.00	U	2.00	P	04/21/2025	17:48	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	17:48	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	17:48	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	17:48	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	17:48	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	17:48	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	17:48	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	17:48	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	17:48	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	17:48	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	17:48	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	17:48	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	17:48	LB135513
CCB04	Aluminum	8.67	+/-20.0	J	20.0	P	04/21/2025	18:21	LB135513
	Antimony	0.11	+/-2.00	J	2.00	P	04/21/2025	18:21	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	18:21	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	18:21	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	18:21	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	18:21	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	18:21	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	500	+/-500	U	500	P	04/21/2025	18:21	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	18:21	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	18:21	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	18:21	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:21	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	18:21	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	18:21	LB135513
CCB05	Aluminum	4.04	+/-20.0	J	20.0	P	04/21/2025	18:49	LB135513
	Antimony	0.11	+/-2.00	J	2.00	P	04/21/2025	18:49	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	18:49	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	18:49	LB135513
	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	18:49	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	18:49	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	18:49	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	18:49	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	18:49	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	18:49	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	18:49	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	18:49	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	18:49	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	18:49	LB135513
CCB06	Aluminum	4.56	+/-20.0	J	20.0	P	04/21/2025	19:14	LB135513
	Antimony	0.11	+/-2.00	J	2.00	P	04/21/2025	19:14	LB135513
	Arsenic	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Barium	10.0	+/-10.0	U	10.0	P	04/21/2025	19:14	LB135513
	Beryllium	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Cadmium	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Calcium	500	+/-500	U	500	P	04/21/2025	19:14	LB135513

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.00	+/-2.00	U	2.00	P	04/21/2025	19:14	LB135513
	Cobalt	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Copper	2.00	+/-2.00	U	2.00	P	04/21/2025	19:14	LB135513
	Iron	50.0	+/-50.0	U	50.0	P	04/21/2025	19:14	LB135513
	Lead	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Magnesium	500	+/-500	U	500	P	04/21/2025	19:14	LB135513
	Manganese	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Nickel	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Potassium	500	+/-500	U	500	P	04/21/2025	19:14	LB135513
	Selenium	5.00	+/-5.00	U	5.00	P	04/21/2025	19:14	LB135513
	Silver	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Sodium	500	+/-500	U	500	P	04/21/2025	19:14	LB135513
	Thallium	1.00	+/-1.00	U	1.00	P	04/21/2025	19:14	LB135513
	Vanadium	5.00	+/-5.00	U	5.00	P	04/21/2025	19:14	LB135513
	Zinc	5.00	+/-5.00	U	5.00	P	04/21/2025	19:14	LB135513

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	20.0	+/-20.0	U	20.0	P	04/29/2025	12:29	LB135595
	Antimony	2.00	+/-2.00	U	2.00	P	04/29/2025	12:29	LB135595
	Arsenic	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Barium	10.0	+/-10.0	U	10.0	P	04/29/2025	12:29	LB135595
	Beryllium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Cadmium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Calcium	500	+/-500	U	500	P	04/29/2025	12:29	LB135595
	Chromium	2.00	+/-2.00	U	2.00	P	04/29/2025	12:29	LB135595
	Cobalt	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Copper	2.00	+/-2.00	U	2.00	P	04/29/2025	12:29	LB135595
	Iron	50.0	+/-50.0	U	50.0	P	04/29/2025	12:29	LB135595
	Lead	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Magnesium	500	+/-500	U	500	P	04/29/2025	12:29	LB135595
	Manganese	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Nickel	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Potassium	500	+/-500	U	500	P	04/29/2025	12:29	LB135595
	Selenium	5.00	+/-5.00	U	5.00	P	04/29/2025	12:29	LB135595
	Silver	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Sodium	500	+/-500	U	500	P	04/29/2025	12:29	LB135595
	Thallium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:29	LB135595
	Vanadium	5.00	+/-5.00	U	5.00	P	04/29/2025	12:29	LB135595
	Zinc	5.00	+/-5.00	U	5.00	P	04/29/2025	12:29	LB135595

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	20.0	+/-20.0	U	20.0	P	04/29/2025	12:44	LB135595
	Antimony	2.00	+/-2.00	U	2.00	P	04/29/2025	12:44	LB135595
	Arsenic	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Barium	10.0	+/-10.0	U	10.0	P	04/29/2025	12:44	LB135595
	Beryllium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Cadmium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Calcium	500	+/-500	U	500	P	04/29/2025	12:44	LB135595
	Chromium	2.00	+/-2.00	U	2.00	P	04/29/2025	12:44	LB135595
	Cobalt	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Copper	2.00	+/-2.00	U	2.00	P	04/29/2025	12:44	LB135595
	Iron	50.0	+/-50.0	U	50.0	P	04/29/2025	12:44	LB135595
	Lead	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Magnesium	500	+/-500	U	500	P	04/29/2025	12:44	LB135595
	Manganese	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Nickel	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Potassium	500	+/-500	U	500	P	04/29/2025	12:44	LB135595
	Selenium	5.00	+/-5.00	U	5.00	P	04/29/2025	12:44	LB135595
	Silver	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Sodium	500	+/-500	U	500	P	04/29/2025	12:44	LB135595
	Thallium	1.00	+/-1.00	U	1.00	P	04/29/2025	12:44	LB135595
	Vanadium	5.00	+/-5.00	U	5.00	P	04/29/2025	12:44	LB135595
	Zinc	5.00	+/-5.00	U	5.00	P	04/29/2025	12:44	LB135595
CCB02	Aluminum	20.0	+/-20.0	U	20.0	P	04/29/2025	13:07	LB135595
	Antimony	2.00	+/-2.00	U	2.00	P	04/29/2025	13:07	LB135595
	Arsenic	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Barium	10.0	+/-10.0	U	10.0	P	04/29/2025	13:07	LB135595
	Beryllium	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Cadmium	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Calcium	500	+/-500	U	500	P	04/29/2025	13:07	LB135595
	Chromium	2.00	+/-2.00	U	2.00	P	04/29/2025	13:07	LB135595
	Cobalt	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Copper	2.00	+/-2.00	U	2.00	P	04/29/2025	13:07	LB135595
	Iron	50.0	+/-50.0	U	50.0	P	04/29/2025	13:07	LB135595
	Lead	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Magnesium	500	+/-500	U	500	P	04/29/2025	13:07	LB135595
	Manganese	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Nickel	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Potassium	500	+/-500	U	500	P	04/29/2025	13:07	LB135595
	Selenium	5.00	+/-5.00	U	5.00	P	04/29/2025	13:07	LB135595

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Sodium	500	+/-500	U	500	P	04/29/2025	13:07	LB135595
	Thallium	1.00	+/-1.00	U	1.00	P	04/29/2025	13:07	LB135595
	Vanadium	5.00	+/-5.00	U	5.00	P	04/29/2025	13:07	LB135595
	Zinc	5.00	+/-5.00	U	5.00	P	04/29/2025	13:07	LB135595