

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** Q1812
Contract: JACO05 **Lab Code:** CHEM **Case No.:** Q1812 **SAS No.:** Q1812
Matrix: Water **Level:** LOW **Client ID:** 285A
Sample ID: Q1824-04 **Spiked ID:** Q1824-04A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	2.41		2.55		500	0		P
Iron	ug/L	75 - 125	31000		13000		25000	72		P
Silver	ug/L	75 - 125	83.9		0.060	J	500	17		P