

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167612BS							
Aluminum	ug/L	10000	10200		102	80 - 120	P
Antimony	ug/L	500	519		104	80 - 120	P
Arsenic	ug/L	500	510		102	80 - 120	P
Barium	ug/L	2500	2590		104	80 - 120	P
Beryllium	ug/L	500	523		105	80 - 120	P
Cadmium	ug/L	500	534		107	80 - 120	P
Calcium	ug/L	50000	51100		102	80 - 120	P
Chromium	ug/L	500	510		102	80 - 120	P
Cobalt	ug/L	500	511		102	80 - 120	P
Copper	ug/L	5000	5020		100	80 - 120	P
Iron	ug/L	25000	27500		110	80 - 120	P
Lead	ug/L	2500	2450		98	80 - 120	P
Magnesium	ug/L	50000	52400		105	80 - 120	P
Manganese	ug/L	5000	5060		101	80 - 120	P
Nickel	ug/L	500	520		104	80 - 120	P
Potassium	ug/L	25000	25500		102	80 - 120	P
Selenium	ug/L	500	502		100	80 - 120	P
Silver	ug/L	500	529		106	80 - 120	P
Sodium	ug/L	50000	53100		106	80 - 120	P
Thallium	ug/L	500	495		99	80 - 120	P
Vanadium	ug/L	500	514		103	80 - 120	P
Zinc	ug/L	5000	4990		100	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: JACOBS Engineering Group, Inc.

SDG No.: Q1812

Contract: JACO05

Lab Code: CHEM

Case No.: Q1812

SAS No.: Q1812

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167634BS Mercury	ug/L	4.0	3.98		100	80 - 120	CV