

Report of Analysis

| | | | |
|--------------------|---|--------------------|-----------------------|
| Client: | JACOBS Engineering Group, Inc. | Date Collected: | 04/15/25 |
| Project: | Former Schlumberger STC PTC Site D3868221 | Date Received: | 04/15/25 |
| Client Sample ID: | RINSE-EB-TANK-041525 | SDG No.: | Q1812 |
| Lab Sample ID: | Q1812-02 | Matrix: | Water |
| Analytical Method: | 8015D DRO | % Solid: | 0 |
| Sample Wt/Vol: | 900 | Units: | mL |
| Soil Aliquot Vol: | | | uL |
| Extraction Type: | | Test: | Diesel Range Organics |
| GPC Factor : | | Injection Volume : | |
| Prep Method : | SW3510 | | |

| | | | | |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| FG015675.D | 1 | 04/16/25 10:05 | 04/16/25 16:42 | PB167614 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|-------------------|-----------------|-------|-----------|----------|------------|---------|
| TARGETS | | | | | | |
| DRO | DRO | 31.0 | J | 7.00 | 56.0 | ug/L |
| SURROGATES | | | | | | |
| 16416-32-3 | Tetracosane-d50 | 16.1 | | 29 - 130 | 80% | SPK: 20 |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit