

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** JACOBS Engineering Group, Inc.
Lab Code: CHEM **Cas No:** Q1812 **SAS No :** Q1812 **SDG No:** Q1812
Matrix Spike - EPA Sample No : PB167614BSD **Datafile:** FG015674.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	158	79	78-117

LCS/LCSD % Recovery RPD : 4.3