

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP041825

SequenceID : BP041825

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓
5. GC/MS Calibration Met:	_____	_____	_____✓
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.  Compound #77 is biased high in the CCC. But not present in the parameters list of the samples analyzed.	_____	_____✓	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

\_\_\_\_\_ ✓ \_\_\_\_\_

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

\_\_\_\_\_ ✓ \_\_\_\_\_

a. B/N Fraction Recovery of a very few compounds are slightly biased high in Q1832-19MS/MSD due to matrix interference. No corrective action is required.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

\_\_\_\_\_ ✓ \_\_\_\_\_

Perylene-d12 internal standard failed in the samples Q1825-09, Q1826-02 which is not associated for required compound.

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_\_ \_\_\_\_\_ ✓ \_\_\_\_\_

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

\_\_\_\_\_ ✓ \_\_\_\_\_

ADDITIONAL COMMENTS:

Recovery of a very few compounds are slightly biased high in PB167625BS/BSD which hit are not present in any associated samples. The data will be used for hardcopies.

Rahul

Analyst

04/18/2025

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: BP041825

SequenceID : BP041825 NA NO YES