

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1812 SAS No.: Q1812 SDG NO.: Q1812

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/17/2025

Lab File ID: BP024320.D Time Analyzed: 10:30

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	303712	7.728	1329210	10.50	880817	14.36
UPPER LIMIT	607424	8.228	2658420	10.998	1761630	14.857
LOWER LIMIT	151856	7.228	664605	9.998	440409	13.857
EPA SAMPLE NO.						
01 RINSE-EB-TANK-041525	219008	7.72	845872	10.50	472895	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.