

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1812 SAS No.: Q1812 SDG NO.: Q1812

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/17/2025

Lab File ID: BP024320.D Time Analyzed: 10:30

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1702910	17.157	1526870	21.609	1367820	24.962
UPPER LIMIT	3405820	17.657	3053740	22.109	2735640	25.462
LOWER LIMIT	851455	16.657	763435	21.109	683910	24.462
EPA SAMPLE NO.						
01 RINSE-EB-TANK-041525	873718	17.15	1001240	21.60	1187980	24.94

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.