

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN041525

Review By	John Carlone	Review On	4/16/2025 9:07:40 AM
Supervise By	Semsettin Yesilyurt	Supervise On	4/16/2025 3:44:21 PM
SubDirectory	VN041525	HP Acquire Method	MSVOA_N HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133665 VP133667,VP133668,VP133669,VP133670,VP133671,VP133672
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP131746 VP133673

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086276.D	15 Apr 2025 10:47		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN086277.D	15 Apr 2025 11:29		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN086278.D	15 Apr 2025 12:21	Comp.#13,16 is on Linear Regression	JC\MD	Ok,M
4	VSTDICC010	VSTDICC010	VN086279.D	15 Apr 2025 12:45	Comp.#43 is on Quadratic Regression	JC\MD	Ok,M
5	VSTDICC020	VSTDICC020	VN086280.D	15 Apr 2025 13:09		JC\MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VN086281.D	15 Apr 2025 13:51		JC\MD	Ok,M
7	VSTDICC100	VSTDICC100	VN086282.D	15 Apr 2025 14:29		JC\MD	Ok,M
8	VIBLK	VIBLK	VN086283.D	15 Apr 2025 15:06		JC\MD	Ok
9	VSTDICV050	ICVVN041525	VN086284.D	15 Apr 2025 15:30		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN041625

Review By	John Carlone	Review On	4/17/2025 9:07:56 AM
Supervise By	Maresh Dadoda	Supervise On	4/17/2025 1:42:06 PM
SubDirectory	VN041625	HP Acquire Method	HP Processing Method 82N041525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133674
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133675,VP133676

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN086285.D	16 Apr 2025 08:38		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN086286.D	16 Apr 2025 09:14	pH#Lot#V12668	JC\MD	Ok,M
3	VN0416MBL01	VN0416MBL01	VN086287.D	16 Apr 2025 09:48		JC\MD	Ok
4	VN0416WBL01	VN0416WBL01	VN086288.D	16 Apr 2025 10:12		JC\MD	Ok
5	VN0416WBS01	VN0416WBS01	VN086289.D	16 Apr 2025 11:52		JC\MD	Ok,M
6	VN0416WBS02	VN0416WBS02	VN086290.D	16 Apr 2025 12:26	Surrogate Fail	JC\MD	Not Ok
7	IBLK	IBLK	VN086291.D	16 Apr 2025 12:50		JC\MD	Ok
8	VN0416WBSD01	VN0416WBSD01	VN086292.D	16 Apr 2025 13:14		JC\MD	Ok,M
9	Q1812-02	RINSE-EB-TANK-0415	VN086293.D	16 Apr 2025 13:38	vial A pH<2	JC\MD	Ok
10	Q1812-03	RINSE-EB-PUMP-0415	VN086294.D	16 Apr 2025 14:02	vial A pH<2	JC\MD	Ok
11	Q1812-01	TB01-041525	VN086295.D	16 Apr 2025 14:26	vial A pH<2 TB	JC\MD	Ok
12	Q1793-02	Storage-Blank-WATER-	VN086296.D	16 Apr 2025 14:50	vial A pH<2	JC\MD	Ok
13	Q1793-03	Storage-Blank-WATER-	VN086297.D	16 Apr 2025 15:14	vial A pH<2	JC\MD	Ok
14	Q1793-04	Storage-Blank-SAMPLE	VN086298.D	16 Apr 2025 15:39	vial A pH<2	JC\MD	Ok
15	Q1793-01	Storage-Blank-SOIL-RE	VN086299.D	16 Apr 2025 16:03	vial A pH<2	JC\MD	Ok
16	Q1760-07MEDL	TP-15-VOCMEDL	VN086300.D	16 Apr 2025 16:27		JC\MD	Ok
17	IBLK	IBLK	VN086301.D	16 Apr 2025 16:51		JC\MD	Ok
18	Q1760-07ME	TP-15-VOCME	VN086302.D	16 Apr 2025 17:15	Need 10X	JC\MD	Dilution

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN041625

Review By	John Carlone	Review On	4/17/2025 9:07:56 AM
Supervise By	Maresh Dadoda	Supervise On	4/17/2025 1:42:06 PM
SubDirectory	VN041625	HP Acquire Method	HP Processing Method 82N041525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133674		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133675,VP133676		

19	IBLK	IBLK	VN086303.D	16 Apr 2025 17:39		JC\MD	Ok
20	VN0416MBS01	VN0416MBS01	VN086304.D	16 Apr 2025 18:03		JC\MD	Ok,M
21	Q1791-14	VB16164	VN086305.D	16 Apr 2025 18:27	vial A pH<2	JC\MD	Ok
22	Q1791-09	VB16193	VN086306.D	16 Apr 2025 18:51	vial A pH<2	JC\MD	Ok
23	Q1791-10	280722	VN086307.D	16 Apr 2025 19:16	vial A pH<2	JC\MD	Ok
24	Q1791-11	280791	VN086308.D	16 Apr 2025 19:40	vial A pH<2	JC\MD	Ok
25	Q1791-13	279312	VN086309.D	16 Apr 2025 20:04	vial A pH<2	JC\MD	Ok
26	Q1791-12	V1207	VN086310.D	16 Apr 2025 20:28	vial A pH<2	JC\MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VN086311.D	16 Apr 2025 20:52	Out of Tune	JC\MD	Not Ok

M : Manual Integration