

Surrogate Summary

SDG No.: **Q1812**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1812-01	TB01-041525	1,2-Dichloroethane-d4	50	50.0	100	70 (74)	130 (125)
		Dibromofluoromethane	50	55.1	110	70 (75)	130 (124)
		Toluene-d8	50	51.3	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.3	103	70 (77)	130 (121)
Q1812-02	RINSE-EB-TANK-041525	1,2-Dichloroethane-d4	50	49.4	99	70 (74)	130 (125)
		Dibromofluoromethane	50	53.6	107	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.3	101	70 (77)	130 (121)
Q1812-03	RINSE-EB-PUMP-041525	1,2-Dichloroethane-d4	50	50.3	101	70 (74)	130 (125)
		Dibromofluoromethane	50	53.7	107	70 (75)	130 (124)
		Toluene-d8	50	51.0	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104	70 (77)	130 (121)
VN0416WBL01	VN0416WBL01	1,2-Dichloroethane-d4	50	49.2	98	70 (74)	130 (125)
		Dibromofluoromethane	50	55.0	110	70 (75)	130 (124)
		Toluene-d8	50	50.6	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100	70 (77)	130 (121)
VN0416WBS01	VN0416WBS01	1,2-Dichloroethane-d4	50	50.6	101	70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101	70 (75)	130 (124)
		Toluene-d8	50	50.8	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.8	100	70 (77)	130 (121)
VN0416WBSD0	VN0416WBSD01	1,2-Dichloroethane-d4	50	53.5	107	70 (74)	130 (125)
		Dibromofluoromethane	50	55.7	111	70 (75)	130 (124)
		Toluene-d8	50	54.6	109	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.6	107	70 (77)	130 (121)

