

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05  
 Lab Code: CHEM Case No.: Q1812 SAS No.: Q1812 SDG NO.: Q1812  
 Lab File ID: VN086286.D Date Analyzed: 04/16/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 09:14  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                      | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD          | 235673        | 8.22  | 430893        | 9.09  | 383563        | 11.87  |
| UPPER LIMIT          | 471346        | 8.724 | 861786        | 9.594 | 767126        | 12.365 |
| LOWER LIMIT          | 117837        | 7.724 | 215447        | 8.594 | 191782        | 11.365 |
| EPA SAMPLE NO.       |               |       |               |       |               |        |
| TB01-041525          | 185374        | 8.22  | 360844        | 9.10  | 336378        | 11.87  |
| RINSE-EB-TANK-041525 | 187101        | 8.22  | 358779        | 9.10  | 327128        | 11.87  |
| RINSE-EB-PUMP-041525 | 180932        | 8.22  | 354597        | 9.10  | 332731        | 11.87  |
| VN0416WBL01          | 215943        | 8.22  | 412690        | 9.10  | 374988        | 11.87  |
| VN0416WBS01          | 282564        | 8.22  | 522491        | 9.09  | 460155        | 11.87  |
| VN0416WBSD01         | 223172        | 8.22  | 407332        | 9.10  | 359125        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.