

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1812 SAS No.: Q1812 SDG NO.: Q1812
 Lab File ID: VN086286.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	177295	13.788				
UPPER LIMIT	354590	14.288				
LOWER LIMIT	88647.5	13.288				
EPA SAMPLE NO.						
TB01-041525	145687	13.79				
RINSE-EB-TANK-041525	145489	13.79				
RINSE-EB-PUMP-041525	151028	13.79				
VN0416WBL01	162383	13.79				
VN0416WBS01	208645	13.79				
VN0416WBSD01	160957	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.