

### Laboratory Control Sample Summary

**Client:** Tully Environmental, Inc

**SDG No.:** Q1815

**Project:** Transfer Station-SPDES

**Run No.:** LB135467

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB135467BS |            |        |                 |            |                 |                     |               |
| TSS       | mg/L       | 550        | 532    |                 | 97         | 1               | 90-110              | 04/17/2025    |