

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1815</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1815</u> sas no.: <u>Q1815</u>
matrix: <u>Water</u>	sample id: <u>Q1804-02</u>	client id: <u>002-35TH-AVE(APR)MS</u>
Percent Solids for Sample: NA	Spiked ID: Q1804-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	142		9.81	J	150	88		P
Iron	ug/L	75 - 125	10200		8870		1500	90		P
Lead	ug/L	75 - 125	446		10.6		500	87		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1815</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1815</u> sas no.: <u>Q1815</u>
matrix: <u>Water</u>	sample id: <u>Q1804-02</u>	client id: <u>002-35TH-AVE(APR)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q1804-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	142		9.81	J	150	88		P
Iron	ug/L	75 - 125	10200		8870		1500	86		P
Lead	ug/L	75 - 125	449		10.6		500	88		P