



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1822

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135473

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Flash Point	o F	82.1	81	101	78-84	04/17/2025