

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1822

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135459

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135459BS							
BOD5	mg/L	198	174		88	1	84.6-115.4	04/16/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1822

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135467

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135467BS							
TSS	mg/L	550	532		97	1	90-110	04/17/2025