

LA135496


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

41272



A= 0.0000e+000

B= 2.4554e-004

C= -4.5344e-002

Rho= 0.9997602

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%I
0.0	0.000	-0.012	-0.012	135	0.000	135					
0.2	0.200	0.219	0.019	1078	0.0 %	1078	0				9
2.5	2.500	2.602	0.102	10781	0.0 %	10781					4
5.0	5.000	4.881	-0.119	20062	0.0 %	20062					-2
7.5	7.500	7.421	-0.079	30409	0.0 %	30409					-1
10.0	10.000	10.089	0.089	41272	0.0 %	41272					1

LB135496 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Apr 2025 09:32:28

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7470A	PPB	21 Apr 2025 10:04:27	S	Std
0.2 - 1	50-2	1078	-	0.2000	7470A	PPB	21 Apr 2025 10:09:27	S	Std
2.5 - 1	52-5	10781	-	2.5000	7470A	PPB	21 Apr 2025 10:11:44	S	Std
5.0 - 1	51	20062	-	5.0000	7470A	PPB	21 Apr 2025 10:14:02	S	Std
7.5 - 1	52-5	30409	-	7.5000	7470A	PPB	21 Apr 2025 10:16:19	S	Std
10.0 - 1	514	41272	-	10.0000	7470A	PPB	21 Apr 2025 10:18:37	S	Std
ICV95 - 1	ICV95	15323	3.7171	-	7470A	PPB	21 Apr 2025 10:21:35	U	SMPL
ICB95 - 1	ICB95	-20	-0.0503	-	7470A	PPB	21 Apr 2025 10:23:49	U	SMPL
CCV01 - 1	CCV01	19570	4.7600	-	7470A	PPB	21 Apr 2025 10:26:06	U	SMPL
CCB01 - 1	CCB01	0	-0.0453	-	7470A	PPB	21 Apr 2025 10:28:22	U	SMPL
CRA - 1	CRA	1000	0.2002	-	7470A	PPB	21 Apr 2025 10:34:36	U	SMPL
HighStd - 1	HighStd	39512	9.6566	-	7470A	PPB	21 Apr 2025 10:36:51	U	SMPL
ChkStd - 1	ChkStd	26983	6.5802	-	7470A	PPB	21 Apr 2025 10:41:50	U	SMPL
PB167657BL - 1	PBW	321	0.0335	-	7470A	PPB	21 Apr 2025 10:48:26	U	SMPL
PB167657BS - 1	LCSW	16608	4.0327	-	7470A	PPB	21 Apr 2025 10:53:15	U	SMPL
Q1821-02 - 1	SOIL-CUTTING-E2	559	0.0919	-	7470A	PPB	21 Apr 2025 10:55:31	U	SMPL
Q1825-03 - 1	TP-9	689	0.1238	-	7470A	PPB	21 Apr 2025 10:57:49	U	SMPL
Q1825-06 - 1	TP-10	5404	1.2816	-	7470A	PPB	21 Apr 2025 11:00:06	U	SMPL
Q1825-09 - 1	TP-11	5089	1.2042	-	7470A	PPB	21 Apr 2025 11:02:23	U	SMPL
Q1826-02 - 1	SOIL-PILE	4897	1.1571	-	7470A	PPB	21 Apr 2025 11:04:38	U	SMPL
CCV02 - 1	CCV02	20181	4.9100	-	7470A	PPB	21 Apr 2025 11:06:53	U	SMPL
CCB02 - 1	CCB02	2	-0.0449	-	7470A	PPB	21 Apr 2025 11:09:08	U	SMPL
Q1826-04 - 1	CONCRETE-PILE	1700	0.3721	-	7470A	PPB	21 Apr 2025 11:11:25	U	SMPL
Q1826-06 - 1	CONCRETE-FLOOR	670	0.1192	-	7470A	PPB	21 Apr 2025 11:13:41	U	SMPL
Q1828-02 - 1	RT4643	89	-0.0235	-	7470A	PPB	21 Apr 2025 11:15:56	U	SMPL
Q1828-04 - 1	RT5417	-19	-0.0500	-	7470A	PPB	21 Apr 2025 11:18:12	U	SMPL
Q1830-10 - 1	SOIL-1	38	-0.0360	-	7470A	PPB	21 Apr 2025 11:20:27	U	SMPL
Q1830-20 - 1	SOIL-2	111	-0.0181	-	7470A	PPB	21 Apr 2025 11:22:43	U	SMPL
Q1832-06 - 1	PL-714-COMP-01	144	-0.0100	-	7470A	PPB	21 Apr 2025 11:25:00	U	SMPL
Q1832-12 - 1	PL-714-COMP-02	129	-0.0137	-	7470A	PPB	21 Apr 2025 11:27:17	U	SMPL
Q1832-18 - 1	PL-714-COMP-03	252	0.0165	-	7470A	PPB	21 Apr 2025 11:29:34	U	SMPL
Q1832-24 - 1	PL-714-COMP-04	2141	0.4804	-	7470A	PPB	21 Apr 2025 11:31:49	U	SMPL
CCV03 - 1	CCV03	19790	4.8140	-	7470A	PPB	21 Apr 2025 11:34:07	U	SMPL
CCB03 - 1	CCB03	-8	-0.0473	-	7470A	PPB	21 Apr 2025 11:36:23	U	SMPL
Q1832-30 - 1	PL-714-COMP-05	1552	0.3357	-	7470A	PPB	21 Apr 2025 11:38:40	U	SMPL
Q1832-36 - 1	PL-714-COMP-6	2068	0.4624	-	7470A	PPB	21 Apr 2025 11:40:56	U	SMPL
Q1832-42 - 1	PL-714-COMP-07	921	0.1808	-	7470A	PPB	21 Apr 2025 11:43:12	U	SMPL
Q1832-42DUP - 1	PL-714-COMP-07DUP	905	0.1769	-	7470A	PPB	21 Apr 2025 11:47:49	U	SMPL
Q1832-42MS - 1	PL-714-COMP-07MS	16171	3.9254	-	7470A	PPB	21 Apr 2025 11:50:04	U	SMPL
Q1832-42MSD - 1	PL-714-COMP-07MSD	16170	3.9251	-	7470A	PPB	21 Apr 2025 11:52:20	U	SMPL
PB167659BL - 1	PBW	103	-0.0201	-	7470A	PPB	21 Apr 2025 11:56:39	U	SMPL
PB167659BS - 1	LCSW	15630	3.7925	-	7470A	PPB	21 Apr 2025 11:58:56	U	SMPL
Q1814-03 - 1	SCRUBBLE-TANK-LIQUID	100	-0.0208	-	7470A	PPB	21 Apr 2025 12:01:13	U	SMPL
Q1814-03DUP - 1	SCRUBBLE-TANK-LIQUIDDUP	296	0.0273	-	7470A	PPB	21 Apr 2025 12:03:30	U	SMPL
CCV04 - 1	CCV04	20699	5.0372	-	7470A	PPB	21 Apr 2025 12:08:27	U	SMPL
CCB04 - 1	CCB04	-45	-0.0564	-	7470A	PPB	21 Apr 2025 12:10:42	U	SMPL
Q1814-03MS - 1	SCRUBBLE-TANK-LIQUIDMS	15494	3.7591	-	7470A	PPB	21 Apr 2025 12:15:17	U	SMPL
Q1814-03MSD - 1	SCRUBBLE-TANK-LIQUIDMSD	15319	3.7162	-	7470A	PPB	21 Apr 2025 12:17:32	U	SMPL
PB167674BL - 1	PBW	-136	-0.0787	-	7470A	PPB	21 Apr 2025 12:19:50	U	SMPL
PB167674BS - 1	LCSW	14973	3.6312	-	7470A	PPB	21 Apr 2025 12:22:07	U	SMPL
Q1822-01 - 1	TW-WTS-06	169	-0.0038	-	7470A	PPB	21 Apr 2025 12:24:24	U	SMPL
Q1822-01DUP - 1	TW-WTS-06DUP	234	0.0121	-	7470A	PPB	21 Apr 2025 12:26:43	U	SMPL
Q1822-01MS - 1	TW-WTS-06MS	14735	3.5728	-	7470A	PPB	21 Apr 2025 12:29:00	U	SMPL
Q1822-01MSD - 1	TW-WTS-06MSD	14597	3.5389	-	7470A	PPB	21 Apr 2025 12:31:17	U	SMPL
Q1841-03 - 1	OILY-WATER	2573	0.5864	-	7470A	PPB	21 Apr 2025 12:33:36	U	SMPL
CCV05 - 1	CCV05	18864	4.5866	-	7470A	PPB	21 Apr 2025 12:39:23	U	SMPL
CCB05 - 1	CCB05	-24	-0.0512	-	7470A	PPB	21 Apr 2025 12:41:38	U	SMPL
PB167621TB - 1	PB167621TB	46	-0.0340	-	7470A	PPB	21 Apr 2025 12:46:13	U	SMPL
PB167622TB - 1	PB167622TB	73	-0.0274	-	7470A	PPB	21 Apr 2025 12:48:28	U	SMPL
Q1832-42LX5 - 1	PB167622TB	195	0.0025	-	7470A	PPB	21 Apr 2025 12:50:44	U	SMPL
Q1832-42A - 1	Q1832-42A	16001	3.8836	-	7470A	PPB	21 Apr 2025 12:52:59	U	SMPL
Q1814-03LX5 - 1	Q1814-03LX5	-131	-0.0775	-	7470A	PPB	21 Apr 2025 12:55:15	U	SMPL
Q1814-03A - 1	Q1814-03A	15390	3.7336	-	7470A	PPB	21 Apr 2025 12:57:32	U	SMPL
Q1822-01LX5 - 1	Q1822-01LX5	-94	-0.0684	-	7470A	PPB	21 Apr 2025 12:59:48	U	SMPL
Q1822-01A - 1	Q1822-01A	17903	4.3506	-	7470A	PPB	21 Apr 2025 13:02:07	U	SMPL
CCV06 - 1	CCV06	18870	4.5881	-	7470A	PPB	21 Apr 2025 13:04:24	U	SMPL
CCB06 - 1	CCB06	-178	-0.0891	-	7470A	PPB	21 Apr 2025 13:06:43	U	SMPL