



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1822
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1822 **SAS No.:** Q1822
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	100	40 - 160	CV	04/21/2025	10:34	LB135496
CRI01	Cadmium	5.99	6.0	100	40 - 160	P	04/21/2025	11:57	LB135512
	Copper	22.1	20.0	110	40 - 160	P	04/21/2025	11:57	LB135512
	Lead	12.6	12.0	105	40 - 160	P	04/21/2025	11:57	LB135512
	Nickel	38.4	40.0	96	40 - 160	P	04/21/2025	11:57	LB135512
	Zinc	40.9	40.0	102	40 - 160	P	04/21/2025	11:57	LB135512