



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	Q1822				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	Q1822	SAS No.:	Q1822		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB95	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	10:23	LB135496	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1822</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1822</u>	SAS No.: <u>Q1822</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	10:28	LB135496
CCB02	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	11:09	LB135496
CCB03	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	11:36	LB135496
CCB04	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	12:10	LB135496
CCB05	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	12:41	LB135496
CCB06	Mercury	0.20	+/-0.20	U	0.20	CV	04/21/2025	13:06	LB135496

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q1822

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1822

SAS No.: Q1822

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	11:53	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	11:53	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	11:53	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	11:53	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	11:53	LB135512

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1822							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1822	SAS No.: Q1822						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	12:23	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	12:23	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	12:23	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	12:23	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	12:23	LB135512
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	13:13	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	13:13	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	13:13	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	13:13	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	13:13	LB135512
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	14:05	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	14:05	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	14:05	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	14:05	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	14:05	LB135512
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	14:55	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	14:55	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	14:55	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	14:55	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	14:55	LB135512
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	15:45	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	15:45	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	15:45	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	15:45	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	15:45	LB135512
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	16:58	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	16:58	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	16:58	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	16:58	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	16:58	LB135512
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	17:49	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	17:49	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	17:49	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	17:49	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	17:49	LB135512
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	18:40	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	18:40	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	18:40	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	18:40	LB135512

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1822
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1822 **SAS No.:** Q1822

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	18:40	LB135512
CCB09	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	19:30	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	19:30	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	19:30	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	19:30	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	19:30	LB135512
CCB10	Cadmium	6.00	+/-6.00	U	6.00	P	04/21/2025	19:47	LB135512
	Copper	20.0	+/-20.0	U	20.0	P	04/21/2025	19:47	LB135512
	Lead	12.0	+/-12.0	U	12.0	P	04/21/2025	19:47	LB135512
	Nickel	40.0	+/-40.0	U	40.0	P	04/21/2025	19:47	LB135512
	Zinc	40.0	+/-40.0	U	40.0	P	04/21/2025	19:47	LB135512