

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1822
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1822 **SAS No.:** Q1822

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167619BS							
Cadmium	ug/L	100	101		101	80 - 120	P
Copper	ug/L	150	157		105	80 - 120	P
Lead	ug/L	500	503		101	80 - 120	P
Nickel	ug/L	250	246		98	80 - 120	P
Zinc	ug/L	100	107		107	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1822
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1822 **SAS No.:** Q1822

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167674BS Mercury	ug/L	4.0	3.63		91	80 - 120	CV