

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 16:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1242-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1242-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1242-5	(5)	5.63	5.63	5.53	5.73	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

Continuing Calib Date: 04/17/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 16:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.77	4.77	4.67	4.87	0.01
Aroclor-1242-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1242-4	(4)	5.04	5.04	4.94	5.14	0.00
Aroclor-1242-5	(5)	5.56	5.56	5.46	5.66	0.00
Tetrachloro-m-xylene		3.69	3.68	3.58	3.78	0.00
Decachlorobiphenyl		8.69	8.68	8.58	8.78	-0.01