

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO110493.D Time Analyzed: 21:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.779	4.679	4.879	496.850	500.000	-0.6
Aroclor-1016-2	4.798	4.698	4.898	494.600	500.000	-1.1
Aroclor-1016-3	4.854	4.755	4.955	481.990	500.000	-3.6
Aroclor-1016-4	4.974	4.875	5.075	485.130	500.000	-3.0
Aroclor-1016-5	5.232	5.132	5.332	487.960	500.000	-2.4
Aroclor-1260-1	6.272	6.172	6.372	428.830	500.000	-14.2
Aroclor-1260-2	6.461	6.360	6.560	409.180	500.000	-18.2
Aroclor-1260-3	6.828	6.729	6.929	401.600	500.000	-19.7
Aroclor-1260-4	7.088	6.988	7.188	400.840	500.000	-19.8
Aroclor-1260-5	7.331	7.230	7.430	388.960	500.000	-22.2
Decachlorobiphenyl	8.732	8.632	8.832	41.680	50.000	-16.6
Tetrachloro-m-xylene	3.688	3.588	3.788	51.360	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/17/2025

Lab Sample No.: AR1660CCC500 Data File : PO110493.D Time Analyzed: 21:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.765	4.666	4.866	493.220	500.000	-1.4
Aroclor-1016-2	4.784	4.684	4.884	486.700	500.000	-2.7
Aroclor-1016-3	4.959	4.860	5.060	481.170	500.000	-3.8
Aroclor-1016-4	5.002	4.902	5.102	447.930	500.000	-10.4
Aroclor-1016-5	5.214	5.115	5.315	496.440	500.000	-0.7
Aroclor-1260-1	6.246	6.146	6.346	436.050	500.000	-12.8
Aroclor-1260-2	6.434	6.334	6.534	428.480	500.000	-14.3
Aroclor-1260-3	6.587	6.487	6.687	421.770	500.000	-15.6
Aroclor-1260-4	7.057	6.957	7.157	414.140	500.000	-17.2
Aroclor-1260-5	7.299	7.197	7.397	407.350	500.000	-18.5
Decachlorobiphenyl	8.685	8.584	8.784	46.330	50.000	-7.3
Tetrachloro-m-xylene	3.685	3.586	3.786	50.300	50.000	0.6