

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO110508.D Time Analyzed: 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.779	4.679	4.879	513.090	500.000	2.6
Aroclor-1016-2	4.798	4.698	4.898	517.490	500.000	3.5
Aroclor-1016-3	4.854	4.755	4.955	502.900	500.000	0.6
Aroclor-1016-4	4.975	4.875	5.075	514.400	500.000	2.9
Aroclor-1016-5	5.231	5.132	5.332	502.540	500.000	0.5
Aroclor-1260-1	6.272	6.172	6.372	507.510	500.000	1.5
Aroclor-1260-2	6.461	6.360	6.560	500.260	500.000	0.1
Aroclor-1260-3	6.828	6.729	6.929	483.980	500.000	-3.2
Aroclor-1260-4	7.088	6.988	7.188	460.460	500.000	-7.9
Aroclor-1260-5	7.330	7.230	7.430	450.860	500.000	-9.8
Decachlorobiphenyl	8.731	8.632	8.832	47.400	50.000	-5.2
Tetrachloro-m-xylene	3.688	3.588	3.788	51.660	50.000	3.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO110508.D Time Analyzed: 03:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.765	4.666	4.866	496.650	500.000	-0.7
Aroclor-1016-2	4.784	4.684	4.884	496.500	500.000	-0.7
Aroclor-1016-3	4.959	4.860	5.060	491.330	500.000	-1.7
Aroclor-1016-4	5.002	4.902	5.102	417.560	500.000	-16.5
Aroclor-1016-5	5.214	5.115	5.315	481.350	500.000	-3.7
Aroclor-1260-1	6.246	6.146	6.346	481.180	500.000	-3.8
Aroclor-1260-2	6.434	6.334	6.534	471.740	500.000	-5.7
Aroclor-1260-3	6.586	6.487	6.687	460.010	500.000	-8.0
Aroclor-1260-4	7.057	6.957	7.157	459.150	500.000	-8.2
Aroclor-1260-5	7.298	7.197	7.397	460.220	500.000	-8.0
Decachlorobiphenyl	8.684	8.584	8.784	50.850	50.000	1.7
Tetrachloro-m-xylene	3.685	3.586	3.786	50.790	50.000	1.6