

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO110517.D Time Analyzed: 08:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.778	4.679	4.879	527.000	500.000	5.4
Aroclor-1016-2	4.797	4.698	4.898	527.560	500.000	5.5
Aroclor-1016-3	4.853	4.755	4.955	506.340	500.000	1.3
Aroclor-1016-4	4.974	4.875	5.075	519.490	500.000	3.9
Aroclor-1016-5	5.230	5.132	5.332	488.740	500.000	-2.3
Aroclor-1260-1	6.270	6.172	6.372	556.430	500.000	11.3
Aroclor-1260-2	6.460	6.360	6.560	526.710	500.000	5.3
Aroclor-1260-3	6.827	6.729	6.929	505.650	500.000	1.1
Aroclor-1260-4	7.087	6.988	7.188	495.160	500.000	-1.0
Aroclor-1260-5	7.329	7.230	7.430	501.060	500.000	0.2
Decachlorobiphenyl	8.731	8.632	8.832	51.730	50.000	3.5
Tetrachloro-m-xylene	3.688	3.588	3.788	52.330	50.000	4.7

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/18/2025

Lab Sample No.: AR1660CCC500 Data File : PO110517.D Time Analyzed: 08:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.764	4.666	4.866	500.510	500.000	0.1
Aroclor-1016-2	4.783	4.684	4.884	501.380	500.000	0.3
Aroclor-1016-3	4.959	4.860	5.060	491.170	500.000	-1.8
Aroclor-1016-4	5.001	4.902	5.102	394.620	500.000	-21.1
Aroclor-1016-5	5.213	5.115	5.315	475.790	500.000	-4.8
Aroclor-1260-1	6.244	6.146	6.346	503.240	500.000	0.6
Aroclor-1260-2	6.432	6.334	6.534	492.170	500.000	-1.6
Aroclor-1260-3	6.585	6.487	6.687	473.950	500.000	-5.2
Aroclor-1260-4	7.056	6.957	7.157	473.420	500.000	-5.3
Aroclor-1260-5	7.296	7.197	7.397	481.020	500.000	-3.8
Decachlorobiphenyl	8.682	8.584	8.784	52.500	50.000	5.0
Tetrachloro-m-xylene	3.685	3.586	3.786	50.860	50.000	1.7