

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM
SubDirectory	BP041425	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024274.D	14 Apr 2025 10:25		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024276.D	14 Apr 2025 11:47	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024277.D	14 Apr 2025 12:27		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024278.D	14 Apr 2025 13:08		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	This calibration is fail for Com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024280.D	14 Apr 2025 15:10	This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024281.D	14 Apr 2025 16:32	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024282.D	14 Apr 2025 17:13	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP041425	BP024283.D	14 Apr 2025 18:35		RC/JU	Ok,M
11	PB167488TB	PB167488TB	BP024284.D	14 Apr 2025 19:16		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041825

Review By	Rahul	Review On	4/21/2025 1:01:03 PM
Supervise By		Supervise On	
SubDirectory	BP041825	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024336.D	18 Apr 2025 10:32		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024337.D	18 Apr 2025 11:13		RC/JU	Ok
3	PB167625BL	PB167625BL	BP024338.D	18 Apr 2025 11:54		RC/JU	Ok
4	PB167625BS	PB167625BS	BP024339.D	18 Apr 2025 12:35		RC/JU	Ok
5	PB167625BSD	PB167625BSD	BP024340.D	18 Apr 2025 13:15		RC/JU	Ok
6	Q1812-03	RINSE-EB-PUMP-0415	BP024341.D	18 Apr 2025 14:00		RC/JU	Ok,NS
7	Q1825-04DL	TP-10DL	BP024342.D	18 Apr 2025 14:41	Dilution not match for compound #84	RC/JU	Not Ok
8	Q1822-01	TW-WTS-06	BP024343.D	18 Apr 2025 15:22		RC/JU	Ok
9	Q1814-01	A3729	BP024344.D	18 Apr 2025 16:03		RC/JU	Ok
10	Q1832-19	PL-714-COMP-04	BP024345.D	18 Apr 2025 16:45		RC/JU	Ok
11	Q1832-19MS	PL-714-COMP-04MS	BP024346.D	18 Apr 2025 17:26		RC/JU	Ok
12	Q1832-19MSD	PL-714-COMP-04MSD	BP024347.D	18 Apr 2025 18:07		RC/JU	Ok
13	Q1828-01	RT4643	BP024348.D	18 Apr 2025 18:48		RC/JU	Ok
14	Q1825-06	TP-10	BP024349.D	18 Apr 2025 19:29		RC/JU	Ok
15	Q1825-09	TP-11	BP024350.D	18 Apr 2025 20:09		RC/JU	Ok
16	Q1826-02	SOIL-PILE	BP024351.D	18 Apr 2025 20:50	Surrogate Fail	RC/JU	ReRun
17	Q1826-04	CONCRETE-PILE	BP024352.D	18 Apr 2025 21:31		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041825

Review By	Rahul	Review On	4/21/2025 1:01:03 PM
Supervise By		Supervise On	
SubDirectory	BP041825	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

18	Q1826-06	CONCRETE-FLOOR	BP024353.D	18 Apr 2025 22:11		RC/JU	Ok
----	----------	----------------	------------	-------------------	--	-------	----

M : Manual Integration