

Surrogate Summary

SW-846

SDG No.: Q1822

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167625BL	PB167625BL	2-Fluorophenol	150	144	96		15 (10)	110 (139)
		Phenol-d6	150	131	88		15 (10)	110 (134)
		Nitrobenzene-d5	100	94.4	94		30 (49)	130 (133)
		2-Fluorobiphenyl	100	92.4	92		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	141	94		15 (44)	110 (137)
		Terphenyl-d14	100	96.6	97		30 (48)	130 (125)
PB167625BS	PB167625BS	2-Fluorophenol	150	131	87		15 (10)	110 (139)
		Phenol-d6	150	121	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	84.3	84		30 (49)	130 (133)
		2-Fluorobiphenyl	100	84.9	85		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	126	84		15 (44)	110 (137)
		Terphenyl-d14	100	89.5	89		30 (48)	130 (125)
PB167625BSD	PB167625BSD	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	122	82		15 (10)	110 (134)
		Nitrobenzene-d5	100	86.7	87		30 (49)	130 (133)
		2-Fluorobiphenyl	100	85.7	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	89		15 (44)	110 (137)
		Terphenyl-d14	100	87.2	87		30 (48)	130 (125)
Q1822-01	TW-WTS-06	2-Fluorophenol	150	50.4	34		15 (10)	110 (139)
		Phenol-d6	150	28.2	19		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.4	77		30 (49)	130 (133)
		2-Fluorobiphenyl	100	73.5	74		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	125	83		15 (44)	110 (137)
		Terphenyl-d14	100	72.9	73		30 (48)	130 (125)